



Workshops of the eighth international brain-computer interface meeting: BCIs: the next frontier

Jane E. Huggins, Dean Krusienski, Mariska J. Vansteensel, Davide Valeriani, Antonia Thelen, Sergey Stavisky, James J.S. Norton, Anton Nijholt, Gernot Müller-Putz, Nataliya Kosmyna, Louis Korczowski, Christoph Kapeller, Christian Herff, Sebastian Halder, Christoph Guger, Moritz Grosse-Wentrup, Robert Gaunt, Aliceson Nicole Dusang, Pierre Clisson, Ricardo Chavarriaga, Charles W. Anderson, Brendan Allison, Tetiana Aksenova & Erik Aarnoutse

To cite this article: Jane E. Huggins, Dean Krusienski, Mariska J. Vansteensel, Davide Valeriani, Antonia Thelen, Sergey Stavisky, James J.S. Norton, Anton Nijholt, Gernot Müller-Putz, Nataliya Kosmyna, Louis Korczowski, Christoph Kapeller, Christian Herff, Sebastian Halder, Christoph Guger, Moritz Grosse-Wentrup, Robert Gaunt, Aliceson Nicole Dusang, Pierre Clisson, Ricardo Chavarriaga, Charles W. Anderson, Brendan Allison, Tetiana Aksenova & Erik Aarnoutse (2022): Workshops of the eighth international brain-computer interface meeting: BCIs: the next frontier, Brain-Computer Interfaces, DOI: [10.1080/2326263X.2021.2009654](https://doi.org/10.1080/2326263X.2021.2009654)

To link to this article: <https://doi.org/10.1080/2326263X.2021.2009654>



Published online: 08 Feb 2022.



Submit your article to this journal [↗](#)



View related articles [↗](#)



View Crossmark data [↗](#)

Workshops of the eighth international brain–computer interface meeting: BCIs: the next frontier

Jane E. Huggins^a, Dean Krusienski^b, Mariska J. Vansteensel^c, Davide Valeriani^d, Antonia Thelen^e, Sergey Stavisky^f, James J.S. Norton^g, Anton Nijholt^h, Gernot Müller-Putzⁱ, Nataliya Kosmyna^j, Louis Korczowski^k, Christoph Kapeller^l, Christian Herff^m, Sebastian Halderⁿ, Christoph Guger^o, Moritz Grosse-Wentrup^p, Robert Gaunt^q, Aliceson Nicole Dusang^{r,s,t,u}, Pierre Clisson^v, Ricardo Chavarriaga^w, Charles W. Anderson^x, Brendan Allison^y, Tetiana Aksenova^z and Erik Aarnoutse^{aa}

^aDepartment of Physical Medicine and Rehabilitation, Department of Biomedical Engineering, Neuroscience Graduate Program, University of Michigan, Ann Arbor, Michigan, USA; ^bDepartment of Biomedical Engineering, Virginia Commonwealth University, Richmond, VA, USA; ^cUMC Utrecht Brain Center, Department of Neurosurgery, University Medical Center Utrecht, The Netherlands; ^dNeurable, Inc; ^eeemagine Medical Imaging Solutions GmbH, Berlin, Germany; ^fUniversity of California, Davis; ^gNational Center for Adaptive Neurotechnologies, US Department of Veterans Affairs, Albany, NY, USA; ^hFaculty EEMCS, University of Twente, Enschede, The Netherlands; ⁱInstitute of Neural Engineering, GrazBCI Lab, Graz University of Technology, Graz, Austria; ^jMedia Lab, E14-548, Massachusetts Institute of Technology (MIT), Cambridge, MA, USA; ^kSiopi.ai, Grenoble, France; ^lG.tec medical engineering GmbH; ^mSchool of Mental Health and Neuroscience, Maastricht University, Maastricht, The Netherlands; ⁿUniversity of Essex, Colchester, UK; ^oG.tec medical engineering GmbH/Guger Technologies OG, Schiedlberg, Austria; ^pResearch Group Neuroinformatics, Faculty of Computer Science, Vienna Cognitive Science Hub, Data Science @ Uni Vienna University of Vienna; ^qRehab Neural Engineering Labs, Department of Physical Medicine and Rehabilitation, Center for the Neural Basis of Cognition, University of Pittsburgh, Pittsburgh, PA, USA; ^rDepartment of Electrical and Computer Engineering, School of Engineering, Brown University; ^sCarney Institute for Brain Science, Brown University, Providence, RI, USA; ^tDepartment of Veterans Affairs Medical Center, Center for Neurorestoration and Neurotechnology, Rehabilitation R&D Service, Providence, RI, USA; ^uCenter for Neurotechnology and Neurorecovery, Neurology, Massachusetts General Hospital, Boston, MA, USA; ^vTimeflux Research Group; ^wIEEE Standards Association Industry Connections group on neurotechnologies for brain-machine interface, Center for Artificial Intelligence, School of Engineering, ZHAW-Zurich University of Applied Sciences, Switzerland; ^xDepartment of Computer Science, Molecular, Cellular and Integrative Neuroscience Program, Colorado State University, Fort Collins, CO, USA; ^yDepartment of Cognitive Science, University of California at San Diego, La Jolla, CA, USA; ^zUniversity Grenoble Alpes, CEA, LETI, Cinatec, Grenoble, France; ^{aa}UMC Utrecht Brain Center, Department of Neurology & Neurosurgery, University Medical Center Utrecht, Utrecht, The Netherlands

ABSTRACT

The Eighth International Brain–Computer Interface (BCI) Meeting was held June 7–9, 2021 in a virtual format. The conference continued the BCI Meeting series' interactive nature with 21 workshops covering the breadth of topics in BCI (also called brain–machine interface) research. Some workshops provided detailed examinations of methods, hardware, or processes. Others focused on BCI applications or user groups. Several workshops continued consensus building efforts designed to create BCI standards and improve comparisons between studies and the potential for meta-analysis and large multi-site clinical trials. Ethical and translational considerations were the primary topic for some workshops or an important secondary consideration for others. The range of BCI applications continues to expand, with more workshops focusing on approaches that can extend beyond the needs of those with physical impairments. This paper summarizes each workshop, provides background information and references for further study, summarizes discussions, and describes the resulting conclusion, challenges, or initiatives.

ARTICLE HISTORY

Received 12 September 2021
Accepted 15 November 2021

KEYWORDS

Brain–computer interface;
brain–machine interface;
neuroprosthetics; conference

1. Introduction

The field of brain–computer interface (BCI) research has many names, most historically originating from related research domains with converging objectives. The terms BCI and brain–machine interface (BMI) are quite common and the term neuroprosthetic also applies. In general, a BCI is a device that interprets information directly from the brain to provide

a means of interacting with technology. Brain activity can be measured using either implanted electrodes or external sensors. The technology can be operated through a variety of methods, including a direct connection between the brain and the effector (e.g. to operate a prosthetic), or a secondary interface such as a keyboard display (e.g. for communication). Recent work has also used electrical stimulation of the brain itself to ‘close the loop’ and provide sensory feedback

about the state of the technology. The defining feature of a BCI is that the brain activity itself is interpreted, the information to control a device is not derived from activity propagated through peripheral nerves. Many BCIs were initially developed for use by people with physical impairments, but the current broad range of applications also targets other neurological and cognitive impairments, abled-bodied users, and even opportunities for human enhancement. The 8th International Brain–Computer Interface Meeting provided a venue for exploration of the breadth of BCI topics and this paper is designed to provide a window into the workshops that occurred at that Meeting.

1.1. The BCI meeting series

The 8th International Brain–Computer Interface Meeting was originally scheduled to be held in 2020. However, due to travel restrictions and health concerns during the global pandemic, the 2020 in-person meeting was postponed to June 7–9th, 2021 and ultimately converted to a virtual meeting format. The goal of the BCI Meeting Series (1999 [1], 2002 [2], 2005 [3], 2010 [4], 2013 [5,6], 2016 [7–9], and 2018 [10,11]) is to create a single venue for people representing all the diverse backgrounds, disciplines, expertise, and application areas necessary for successful and practical BCI research and development.

The Eighth International Brain–Computer Interface (BCI) Meeting was hosted in the Pheedloop platform (Toronto, Ontario, Canada), which managed individual sessions using the Zoom platform (San Jose, California, USA). Poster sessions and social events were held on the GatherTown platform (gather.town). This Meeting was attended by 395 delegates from 35 countries, a significant growth from the 50 delegates in 1999 [1], although not quite as many as the previous in-person meeting in 2018. Respondents to the 2021 BCI Meeting evaluation survey identified themselves as 40% students, 13% postdocs, 25% faculty members, and 22% other. The BCI Meeting Series is intentionally designed to promote interaction between different groups and different career stages and has advanced the careers of numerous BCI researchers. Many activities are designed to provide educational content and networking opportunities for students and early-career investigators. The 2021 BCI Meeting had a theme of ‘BCIs: The Next Frontier’. The workshops of the BCI Meeting Series provide examples of how BCIs are advancing the frontiers of science and details on both how close we are to realizing new applications and the challenges that remain to be overcome. The

workshop summaries presented here serve as an overview of the current status of BCI research and development and present a roadmap to the next steps needed to advance that frontier.

1.2. Organization of workshop summaries

Workshops for the BCI Meetings are proposed by members of the BCI community, then evaluated and curated by the Program Committee. For the virtual BCI Meeting of 2021, the workshops were assigned to four different schedule slots with three to four workshops running concurrently. In addition, six of the workshops volunteered to run as part of a five-month preliminary series of ‘BCI Thursdays’. These workshops were the same length and format as the workshops that occurred during the Meeting but did not overlap with other BCI Society events and had a separate registration structure. However, they retained the strong emphasis on attendee participation that is central to workshops of the BCI Meeting series. The BCI Thursday series also included free events designed to provide technical background for students on cutting-edge topics in BCI research.

The workshop summaries presented here are divided into three themes and ordered to provide a progression of topics. They can be read sequentially as an overview of the field or separately to provide details on a topic of interest. However, acronyms are only defined on their first use. For each summary, we report the primary organizer, who is also a co-author of this paper, and list all additional presenters. Each summary is designed to introduce the workshop topic, the latest developments or central ideas presented in the workshop, and the topics of discussion and eventual conclusions. Of course, nothing will substitute for the actual experience of being part of an interactive workshop, even a workshop in a virtual platform. However, the summaries are intended to at least provide an overview and pointers to the information that workshop attendance would have provided. Further, the summaries provide the key points, conclusions, or consensus opinions that resulted from the workshop discussions and may include opportunities to participate in ongoing discussions or collaborations.

Each workshop focused on a specific topic area, yet these topics overlap and complement each other, so that the summaries sometimes create a mosaic examining related ideas from different angles and at other times build on each other. For example, the workshops ‘*Toward an international consensus on user characterization and BCI outcomes in settings of daily living*’ and ‘*On the need of good practices and standards for Benchmarking Brain–Machine Interfaces*’ examine

different aspects of standards. Similarly, BCI use for children and people with congenital disabilities are examined in the pair of workshops ‘*The design of effective BCIs for children*’ and ‘*Non-invasive BCIs for people with cerebral palsy*’.

Three general themes provide the structure for this article, although many alternative organizations could be proposed. The themes are independent of the time slot in which the workshop occurred. The first theme is Tools and Methods and contains workshops providing detailed examination of a particular hardware, software, or analysis method. The second theme is BCIs for Specific Populations or Applications and is less concerned with hardware and software than with the outcome produced or the common considerations for working with a specific group. The final theme is Expanding BCI Usability and Availability. The workshops in this theme focus on big picture topics such as standards, translational issues, and ethics as well as the expansion of BCIs into the broad consumer market through applications such as entertainment and human enhancement.

The trajectory of these three themes, and the workshop summaries presented here, creates a progression from foundational topics to translational efforts for standardized clinical applications and BCIs for the population at large. Together these workshops show the diversity of BCI applications and intended users and the complexity of the issues that must be solved to make BCIs into useful tools for the many intended user groups.

2. Tools and methods

2.1. Focal bi-directional brain computer interfacing with concentric electrode technology

Organizer: Charles Anderson (Colorado State University)

Additional Presenters: Walter Besio (University of Rhode Island and CREMedical), Barry Oken (Oregon Health & Science University), Myles McLaughlin (KU Leuven)

This workshop focused on EEG BCI experiments and stimulation studies using tripolar concentric-ring electrodes (TCREs) and the advantages of this technology over conventional disc electrodes. Compared to conventional disc electrodes, TCREs have significantly better spatial resolution and signal-to-noise ratio [12–14]. TCREs increase signal bandwidth for high-frequency signals useful for localizing epileptic brain regions and possibly imagined movements [15,16]. Imagined movement BCI improved significantly with TCREs [17,18].

TCREs’ increased spatial resolution and signal-to-noise ratio may enable discrimination between finger movements, currently only possible with implanted electrodes. Experiments involving real and imagined finger movements found that EEG from TCREs produced significantly better discrimination among movements of individual fingers (about 70% correct classification) than conventional disc electrodes (about 40%) [19].

TCREs are safe for stimulation [20,21] and can be used for seizure control [22–26]. The stimulation can block epileptogenesis [27] and alter neurotransmitters to increase the effectiveness of anti-seizure drugs [28–30]. Stimulation experiments are underway to determine if transcranial focal stimulation via concentric ring electrodes is effective for modulating human brains.

Pain is a common medical problem but difficult to objectify as a personal experience of a sensation. Using TCREs both to selectively stimulate pain fibers and to record pain-related evoked potentials (PREPs) is one method of objectifying pain sensation [31–37]. Custom-made concentric stimulating electrodes can selectively stimulate pain afferents where conventional electrical stimulation with mono- or bi-polar stimulating electrodes failed. TCREs delivered paired electrical stimulations to the dorsal non-dominant hand. PREPs were recorded at Cz referenced to ear. For control participants, average PREP N1-P2 amplitude was significantly diminished by electroacupuncture. In another experiment control participants showed the expected habituation of PREP N1-P2 amplitude over time, but those with chronic low back pain showed an increase in PREP amplitude, presumably a physiological marker of central sensitization, the increased responsiveness to sensory information such as nociception.

TCREs on the skull under the skin may be an effective middle ground between implanted stimulation electrodes and the non-invasive but less effective transcranial stimulation. TCREs provide higher magnitude stimulation in gray and white matter than transcranial stimulation. Focused and unfocused stimulation on neurons have been studied in Macaque. Increased spatial precision with TCREs was demonstrated when stimulating rat motor cortex area for rear limb movement. Conventional electrodes produced movement in both contralateral and ipsilateral limbs, but TCREs only produced contralateral limb movement [38].

Discussion covered practical considerations and design variations, including different numbers of rings and different spacing. TCREs sizes include 10 mm, 6 mm, 4 mm, and even 3.5 mm. TCREs use 10–20 paste, but work on using gels and possible dry electrode

designs are being considered. Caps to hold TCRES were described, but need work for the smallest TCRES. Two disadvantages of TCRES are the need for a custom pre-amplifier from CREMedical and for precise scalp placement because higher spatial precision means steep attenuation over short distances. Laplacian transforms can be applied to EEG recorded from conventional disc electrodes, but 92 disc electrodes are required to obtain results similar to that provided by one TCRES. Publicly available sample data recorded from TCRES can be found at <https://www.cs.colostate.edu/~anderson/res/eeg/tripolar/tripolar.zip>.

2.2. Invasive brain computer interface technology: open loop and closed loop decoding applications

Organizer: Christoph Kapeller (g.tec medical engineering GmbH, Austria)

Additional Presenters: Kyousuke Kamada, MD, PhD, (Megumino Hospital, Japan); Aysegul Gunduz, PhD, (University of Florida, USA); Peter Brunner, PhD, (Washington School of Medicine, St. Louis, USA); Kai Miller, MD, PhD (Mayo Clinic Rochester, Minnesota, USA).

The workshop discussed state-of-the art BCI applications using open-loop and closed-loop decoding and neuromodulation. Implementation of these experimental setups in existing BCI platforms was also discussed.

Invasive electroencephalographic (iEEG) signals, such as electrocorticography (ECoG) or stereo EEG, contain information with high spatial and temporal resolution [39]. Several invasive BCIs have been realized over the past two decades. Closed-loop invasive BCIs have been used for control of prosthetic limbs [40] as well as avatars or cursors [41,42]. Open-loop invasive BCIs have been used for decoding of speech [43–46], movements [47,48] and vision [49,50]. Establishing useful invasive BCI applications requires interdisciplinary efforts for the development of sensors and machine learning algorithms, with specialized efforts to make the resulting technology practical for a medical environment and matched to each individual's clinical indications. Further, the risk of implanting sensors has to be surpassed by the benefit that the BCI provides to meet the specific need of each patient [51].

Recent developments showed a transition from proof-of-concept demonstrations to clinical applications, including open-loop decoding for brain mapping [52–54] and BCI implants [55]. Such implants can provide ALS patients with a powerful BCI [42] and will be further investigated over the next years. The concept of open-loop electrical brain stimulation for neuromodulation has been widely used in presurgical brain

mapping. Stimulating the somatosensory cortex can induce sensation in individual fingers [56], while stimulating the visual cortex causes illusory percepts like appearing faces or moving rainbows [57]. Open-loop deep brain stimulation (DBS) has been utilized for more than 40 years to manage tremor [58]. More recently, DBS has been used to treat Parkinson's disease, Tourette syndrome, dystonia, and depression [59]. Closed-loop stimulation based on iEEG signals improves the battery lifetime during the treatment of Tourette syndrome [60] and essential tremor [61]. Most of the aforementioned studies required the integration of sensors and amplifiers into signal processing platforms that are capable of real-time processing and synchronized with the patient's condition and/or stimulus presentation. Example BCI platforms in the workshop were BCI2000 [62] and the rapid prototyping platform g.HISys in MATLAB/Simulink [63].

2.3. Riemannian geometry methods for EEG preprocessing, analysis and classification

Organizer: Louis Korczowski (Siopi.ai)

Additional Presenters: Marco Congedo (GIPSA-lab, CNRS, Université Grenoble- Alpes), Florian Yger (LMSADE, CNRS, Univ. Paris-Dauphine, PSL Research Univ.), Sylvain Chevallier (LISV – UVSQ – Univ. Paris-Saclay), Pierre Clisson (Timeflux Research Group), Quentin Barthélemy (Foxstream)

Riemannian Geometry (RG) is a subject of growing interest within the BCI community. Machine learning methods based on RG have demonstrated robustness, accuracy, and transfer learning capabilities for the classification of motor imagery [64], ERPs [65], SSVEPs [66], sleep stages [67], and other mental states [68]. This workshop provided an overview of RG, demonstrating its practical use for signal pre-processing, data analysis, mental state classification, and regression.

RG was first applied to BCI in 2010 [64]. Key articles highlighting different applications of RG include multi-class classification (e.g. minimum distance-to-mean (MDM) classifier) [69], transfer learning (e.g. Riemannian Procrustes Analysis) [70,71], the first online BCI system using it (e.g. Brain Invaders) [65,72], and milestone-like performance of RG methods in international competitions [73,74]. Intrinsic properties of RG methods were discussed to explain their performances (e.g. simple parametrization of models, robustness induced by affine-invariant metrics) but also some drawbacks and how they can be managed (e.g. sensitivity to rank deficiency at high dimensionality) [75,76]. Interestingly, RG can be used in combination with other effective methods such as common-

spatial pattern and/or deep learning to outperform methods using Euclidean space alone, e.g. by projecting data in a tangent space [74].

The ecosystem of open-source libraries (that was scattered and scarce before) is now mature enough to improve several steps of the BCI system. For example, Riemannian methods outperforms Euclidean methods in accuracy and simplicity in use cases such as automatic artifact detection (e.g. Riemannian potato) [77,78] or ERP classification (e.g. MDM with super covariance matrix). These performances are tested using the fair benchmarking approach [79] and are easy to replicate in online BCI thanks to libraries such as Timeflux [80].

Despite its performance advantages, publication data from <https://www.dimensions.ai/> show that articles mentioning new contribution of ‘Riemannian Geometry’ applied to BCI has remained in the range of 7 to 21 per year in the period 2016–2020 (mean citations: 27.71). For comparison, mention of ‘common-spatial patterns’ associated with BCI increased from 71 to 119 articles per year (mean citations: 20.75) and ‘deep learning’ from 15 to 179 articles per year (mean citations: 11.67) in the same period.

We argue that the gap between the observed performance of RG applied to BCI and the proposal number of contributions in this field may be attributed to some combination of a perceived lack of easily accessible resources to make RG widely available to BCI research (e.g. 65.7% of respondents to the workshop questionnaire had never used RG before) and the lack of reproducible tools for benchmarking different methods while taking into consideration datasets heterogeneity (discussed at the previous BCI meeting workshop [81]).

This workshop was created to address these issues by increasing awareness of available resources for RG and encourage benchmarking with tools such as MOABB on a larger scale of datasets [79]. We encourage everyone to report benchmarking results. Further, we invite everyone to join us by using the open-source RG tools, and by contributing to the improvement of these tools either by providing feedback, or contributing to the open-source project pyriemann. All the workshop resources are accessible, including slides, code tutorial, online demo, exhaustive workshop Q&A, and linked data: <https://github.com/lkorcowski/BCI-2021-Riemannian-Geometry-workshop>.

2.4. Open-source python tools for BCIs

Organizer: Pierre Clisson (Timeflux Research Group)

Additional Presenters: Raphaëlle Bertrand-Lalo (Timeflux Research Group), Sylvain Chevallier (LISV, Université Paris-Saclay), Marco Congedo (GIPSA-lab, CNRS, Université Grenoble-Alpes)

Python started as a general-purpose programming language but has evolved into a tool of choice for the scientific community, quickly overtaking specialized languages such as R and MATLAB [82]. Several factors account for its success: Python is easy to learn, has a strong community, and benefits from a rich and efficient data science ecosystem.

This workshop had a two-fold objective: give an overview of the Python BCI landscape and provide hands-on instructions on a few chosen open-source tools.

As a foundation for the focus on practical BCI, we first reviewed the main BCI paradigms and the typical workflow of a BCI pipeline. We discussed common challenges for BCI applications: the need for precise synchronization of the EEG signal and the stimuli, the difficulty of obtaining good quality signals in real-life conditions, and the challenges of calibration.

Riemannian geometry (RG) for EEG-based BCI [65,83] has produced state-of-the-art results in international competitions [76]. Machine-learning algorithms based on RG offer many advantages. They are computationally efficient and thus suitable for online applications. They usually converge to optimal results relatively quickly, reducing calibration duration (ongoing studies on transfer learning are attempting to remove this phase completely [70,84]). Finally, they do not depend on the BCI paradigm and work equally well for ERP, SSVEP, and motor imagery tasks.

PyRiemann [85] is an actively maintained Python package for manipulating covariance matrices. It implements multiple data transformation techniques and classification methods. Workshop participants were guided through a Python notebook and instructed on using this library with concrete examples.

The RG framework includes multiple signal classification strategies and BCI researchers use many other algorithms, such as Logistic Regression, Regulated LDA, Support Vector Machines, and Neural Networks [86]. Valid comparisons between methods are essential. The Mother Of All BCI Benchmarks (MOABB) [79,87] project offers comprehensive comparison tools that enable ranking new and existing algorithms with publicly available datasets, paving the way for reproducible research. We reviewed a practical example and explained the underlying code.

Timeflux (<https://timeflux.io/>) [80] is an open-source framework for building online BCIs. It is capable of acquiring, recording, and processing biosignals in real-time. It can also present precisely scheduled stimuli. It works hand-in-hand with PyRiemann and MOABB and rests on the shoulders of standard libraries such as Pandas [88], Scikit-learn [89], Lab Streaming Layer [90], and HDF5 [91]. It comes with a rich set of nodes

and plugins for dynamic epoching, matrix manipulation, digital signal processing, machine learning, and other tools. It also provides a convenient JavaScript API for developing web interfaces. We reviewed the architectural principles of Timeflux and explained how to use it to design a P300 speller, finishing with a functional demo that runs in a web browser.

We only introduced the potential of the Python language for the BCI field. For instance, we only briefly described MNE [92], a full-fledged framework for offline analysis of EEG and MEG signals. This workshop provided a good starting point for further exploration. The presentation slides, notebooks, and code are publicly available [93].

2.5. Artificial intelligence in brain-computer interfacing

Organizer: Moritz Grosse-Wentrup (University of Vienna)

Additional Presenters: Tonio Ball (University of Freiburg), Aldo Faisal (Imperial College London), Gernot Müller-Putz (Graz University of Technology)

Artificial intelligence (AI) methods in general, and deep learning algorithms in particular, have revolutionized the field of machine learning [94]. Current AI systems outperform human experts in various cognitively challenging tasks [95,96] and have enabled scientific insights that arguably could not have been obtained by human intelligence alone [97]. More recently, deep learning methods have been adapted to and developed for brain decoding and BCI systems [98,99]. Building on a long history of discussions on the benefits of nonlinear decoding methods in BCI [100], this workshop discussed whether AI can outperform traditional BCI machine learning methods and which challenges should be addressed to realize the full potential of AI in BCI.

The consensus on the current performance of AI-BCI methods was that they perform essentially on par with the best non-deep decoding algorithms. However, a rigorous comparison of state-of-the-art Riemannian decoding methods [76,101] with AI algorithms has yet to be done. The workshop participants concluded that a large-scale brain decoding challenge, e.g. hosted by a major AI or machine learning conference, would be well suited for realizing a fair comparison of competing decoding architectures (e.g. <https://beetl.ai/>).

The workshop participants then considered which issues prevent, at least so far, AI methods from revolutionizing BCI systems in the same way they have already transformed other data-driven applications. The primary bottleneck identified in the discussion was the absence of large-scale datasets in the field of BCI.

These datasets would ideally comprise thousands or even millions of BCI users from heterogeneous settings, i.e. including numerous experimental paradigms, recording setups, and user groups. While the workshop participants acknowledged the efforts of the BCI community to record large-scale datasets [102], they also noted that collecting datasets on a similar scale as those available in other scientific disciplines [103] is probably beyond the capabilities of the academic community. Consequently, the discussion shifted to the role of commercial BCI applications in recording and providing access to large-scale datasets. Several consumer EEG headsets have reached market readiness with the expectation of prompt deployment in passive BCI applications [104]. Comprehensive access to data recorded by these applications could provide the large-scale datasets required to realize the full potential of AI-BCI systems. In particular, the heterogeneous nature of such data, which stands in contrast to the homogeneous data typically recorded in academic settings, could be considered an advantage. The diversity of data might be leveraged to create feature representations that are user- as well as hardware-independent. Such feature representations would be essential to realize zero-training BCIs for commercial applications [105–107].

However, leveraging commercially recorded EEG datasets poses significant practical, legal, and ethical challenges. It is unclear what incentives companies would have to share their data publicly. Also, procedures would have to be developed that realize informed consent and honor data privacy regulations. The workshop participants considered an active engagement of the BCI community with industrial partners essential to make large-scale datasets a reality and realize the full potential of AI-BCI systems.

2.6. Adaptation in closed-loop BCIs

Organizer: Tetiana Aksenova (University Grenoble Alpes, CEA, LETI, CLIMATEC)

Additional Presenters: Amy L. Orsborn (University of Washington), Martin Bogdan, Sophie Adama (Universität Leipzig), Blaise Yvert (U1205 Inserm, University Grenoble Alpes), José del R. Millán (University of Texas at Austin), Jean Faber (Universidade Federal de São Paulo)

BCI decoders calibrated in an open-loop, offline paradigm but then applied in close-loop, online paradigm show a significant drop in decoding performance. Adaptive algorithms in a closed-loop session decrease this shortcoming by directly adjusting BCI parameters to incoming data. In addition, both the user and machine learn in a closed-loop BCI.

Closed-loop paradigms are often applied to BCIs that decode motor signals. Intracranial ECoG [108,109] from a participant with tetraplegia was decoded with a fully adaptive decoder to operate a 4-limb exoskeleton. The decoder used an adaptive Markov mixture of multi-linear experts [110] to switch between independent decoders (experts) to interpret multiple degrees of freedom.

Closed-loop paradigms enable user/decoder co-adaptation to maximize performance through synergistic user-machine interactions between the two learners e.g [111]. However, learning trajectory models are needed to optimize these co-adaptive systems. A new game-theoretic model of co-adaptation [112] provides a framework to analyze system equilibria and predicts learning trajectories, but requires validation.

The balance of decoder vs patient adaptation is important. EEG-based motor BCIs illustrate the pros and cons of extensive machine-learning adaptation. Non-supervised context-aware algorithms can rapidly adapt so users can use a language model-based speller [113] without a calibration phase [114,115]. However, this does not promote user learning – EEG patterns for BCI commands actually became less separable with practice rather than improving [115]. True mutual learning, where decoder and user learn from each other, seems to require slow decoder adaptation to promote improved EEG features [116] as seen in several longitudinal studies [117].

Mutual learning implies cortical plasticity and the BCI use as a neurorehabilitation tool specifically designed to support plasticity (i.e. user learning). A clinical trial in patients with severe hand plegia from stroke compared the effect of BCI-operated vs random functional electrical stimulation. Only the BCI group had significant and clinically important functional improvement and a significant increase of functional connectivity in the damaged sensorimotor hemisphere [118]. Regulation of the magnitude of the required EEG response was critical to keep the patient's attention high and promote recovery.

Hybrid BCIs (HBCIs) integrate brain and non-brain data sources with different classifiers schemes (serial, parallel, mixed) to achieve better results [119]. Thus, neuroplasticity can happen in multiple dimensions and temporal scales. Different learning times are associated with different physiological systems such as autonomic learning (heart/breath adaptation) [120,121], motor learning (agency and control refinement) [122,123], central learning (cortical adaptations) [124], and cognitive learning (embodiment, ownership, and spatial

perception) [125]. HBCIs therefore present a more complex challenge for balancing classifier adaptation rate vs. neural plasticity.

Adaptive BCIs also exist for non-motor applications. The hybrid Adaptive Decision-Making system was designed for a patient with complete locked-in syndrome (CLIS) and uses multiple EEG features (Granger causality, the imaginary part of the coherency, and multiscale sample entropy) to increase the probability of correctly evaluating consciousness level [126]. Caregiver observations regarding the patient's state were input into the machine learning system to personalized consciousness-level estimation. An adaptive speech BCI application illustrates the risk of audio contamination of neuronal activity recordings [127].

Group discussion placed a priority on developing better understanding of co-adaptation from both theoretical and experimental viewpoints to optimize BCI training and user benefit.

2.7. Optimising BCI performance by integrating information on the user's internal state

Organizer: Sebastian Halder (University of Essex)

Additional Presenters: Philipp Ziebell, University of Würzburg), Angela Riccio (Fondazione Santa Lucia), Yiyuan Han (University of Essex)

Ideally, a BCI could detect the physical and mental state of the user and adapt accordingly to allow optimal BCI control for both unimpaired and motor impaired end-users. This adaptation could (1) determine when to start, pause or stop a BCI session, (2) adapt parameters of the BCI session such as trial length, stimulus, and feedback modality, or (3) switch between BCI and other assistive technology types. User-centered design (UCD) is critical to optimize BCI control in this manner [128]. In general terms, an assistive technology should enable a person with a disability to overcome barriers in daily life, education, work, or leisure [129]. This can only be achieved if the needs and requirements of the user are investigated [130,131]. Regarding BCI design, the cognitive [132–134] and physical [135,136] characteristics of end-users need to be considered [132,133]. Based on this knowledge, we can implement a system that adapts to the internal state of the user.

The UCD evaluation process is built around metrics to determine effectiveness (accuracy in percent of correct responses), efficiency (information transfer rate in bits/min and subjective workload) and satisfaction (via visual analog scale, questionnaire, or user interview) [137,138]. These metrics should also inform earlier stage BCI development before end-user evaluation [139,140]. Further factors should be considered when

designing the BCI paradigm, for instance, the design of tasks, feedback, instructions, and signal processing [86,141–143]. Performance may improve via engaging task design (e.g. a ‘Star Wars Mission’ task) and exploring different stimulus modalities (such as auditory and tactile) and better understanding of the mechanisms underlying training with a BCI [140,144].

User characteristics ranging from physiological (e.g. the amplitude of the sensorimotor rhythm during rest [145]) to psychological (e.g. the ability to concentrate [132,146]) can influence performance in varying degrees. For example, a user with a traumatic brain injury may be in a minimally conscious state with only transient windows of consciousness [147,148]. Identifying such windows is an undeniable prerequisite to BCI control [149]. Evaluation of the efficacy of such measures and any new measures that will be developed can be accomplished during pharmacologically induced loss of consciousness such as the Wada test [150]. More subtle influences on BCI control may arise due to mood and motivation, fatigue and workload or whether the user is experiencing pain, which can be detected using integrative features such as phase-based connectivity [151–153]. Ideally, the BCI could adapt to all changes in the users’ state. Doing this efficiently requires knowledge of features in the EEG (or other signals) that reflect the state of the user.

Many challenges must be resolved before the full potential of the state of the user can be reliably used to optimize BCI performance. The main challenge comes from the variety of states that need to be decoded, each requiring the identification of signal features that reflect these states, and integrating real-time identification of the states into the BCI design and usage environment.

3. BCIs for specific populations or applications

3.1. The design of effective BCIs for children

Organizers: James J.S. Norton (National Center for Adaptive Neurotechnologies), Disha Gupta (National Center for Adaptive Neurotechnologies), Eli Kinney-Lang (University of Calgary)

Additional Presenters: Kim Adams (University of Alberta), Tom Chau (University of Toronto), Erica Floreani (University of Calgary), Kathleen M. Friel (Burke Neurological Institute), Dion Kelly (University of Calgary), Adam Kirton (University of Calgary), Ilyas Sadybekov (University of Calgary), Corinne Tuck (Glenrose Rehabilitation Hospital-I CAN Centre).

BCIs have the potential to enhance, restore, or replace function in children with neurodevelopmental disorders, neurodegenerative disorders, and severe

motor disabilities caused by stroke, spinal cord injury, or other acquired injuries [154–157]. However, few studies have investigated BCIs for children [158–161] and these studies show conflicting results; it remains unclear whether children – especially those with neurological disabilities – can effectively use BCIs. Thus, this workshop was organized into three discussion panels that:

- (1) **Examined how BCIs can improve children’s quality-of-life** – Children can use BCIs to [162] communicate, play games, and express themselves creatively. The greatest benefit BCIs offer children with motor disabilities is a sense of control, motivating children to engage more with BCIs and enabling them to practice repetitive tasks that lead to learning. Thus, the child’s perception of a successful BCI may not match that of a researcher. For example, operating a BCI using a combination of brain activity and artifacts may improve the child’s life and be considered a success from the child’s perspective. Therefore, special consideration is needed to simultaneously engage children in activities that are educational, therapeutic, meet the goals of researchers, and are engaging for the children. Recommended strategies are gamification [163–167] and close interdisciplinary collaboration between diverse experts.
- (2) **Discussed the interfacing, signal-processing, and physiological challenges encountered during the design of BCIs for kids** – Developing BCIs for children presents unique signal acquisition, data analysis, and reporting challenges [154]. Signal acquisition hardware for pediatric BCIs needs to be more portable, lighter, more comfortable, and easier to use (e.g. faster setup, dry electrodes, robust to artifacts). Presently only a few signal analysis pipelines exist for pediatric BCIs [168,169], due in part to differences in the EEG from children compared to adults [170]. For example, P300 timing varies more in children and BCIs may be more fatiguing for children. Improved and consistent reporting of demographic information and experimental details would allow for better cross-study analyses. Lastly, improved user interfaces are an area of critical need for pediatric BCIs.
- (3) **Considered the use of BCIs for children as augmentative and alternative communication devices and for rehabilitation in clinical settings** – The design of BCIs for communication and rehabilitation in children benefits from a patient-

centered and neurologic deficit-specific approach [161,171]. For example, many children express an interest in using BCIs for gaming and social play. Collaborative and competitive interactions between family members, and especially siblings, are a critical social outlet for children with motor deficits that motivate them to use BCIs. Neurological deficits may be caused by damage to small areas of the brain that were acquired very early in life. Thus, the brain may reorganize and researchers should work with clinicians to consider neuroplasticity in the design of BCIs for children [172,173]. In addition, working with clinicians and families will increase awareness of the potential of BCIs for children [174].

As members of the pediatric BCI community, we must put children first, understand what children want out of BCIs, and make it happen.

3.2. Non-invasive BCIs for people with cerebral palsy

Organizer: Jane E. Huggins (University of Michigan)

Additional Presenters: Katya Hill (University of Pittsburgh), Petra Karlsson (Cerebral Palsy Alliance, University of Sydney), Reinhold Scherer (University of Essex)

This workshop included extensive discussion about BCI design considerations for people with cerebral palsy (CP), the most common childhood physical disability [175]. CP is caused by injury or genetic abnormalities affecting the brain early in life leading to 15–19% without a communication method even with assistive technology [176–179]. However, BCIs that provide augmentative and alternative communication (AAC) for individuals with adult-onset impairments may unintentionally rely on skills that people with CP have not had an opportunity to learn.

Issues from the workshop *Design of Effective BCIs for Children* apply to children and adults with CP because of missed educational opportunities. Even those who have successful communication technology may need a BCI as age increases the severity of motor impairments. This makes BCI a competitive access option. For example, a participant with CP had similar communication rates on an AAC device with head-pointer access (1.33 words-per-minute, wpm) and BCI access (1.29 wpm).

Overall, BCI studies with people with CP show mixed results [162,180,181]. Some comparisons of BCI designs showed that SSVEP and SMR designs were preferred to the P300 design and had better performance [181].

Other comparisons of naïve users showed that some had significant SMR-BCI control (2 classes, $82 \pm 12\%$), others significant SSVEP-BCI control (4 classes, $43 \pm 7\%$), but few could use both and some could not use any BCI [182,183].

Such results raise the specter that current BCI methods may not be appropriate for people with CP. If a person has no voluntary motor control, can they operate a motor imagery BCI? Can people with limited access to schooling count flashes of a P300 BCI or perform mental arithmetic or spatial navigation?

EEG recordings are complicated in people with CP due to head shape variations or improper electrode cap fit [184,185] as head asymmetry is reported among 40% of people with the most severe impairments from CP [186] and microcephaly at 30% [187] to 60% [188]. Abnormal neuroanatomy can also cause unusual localization of cortical function [189]. The impact on BCI is uncertain, but people with severe CP can benefit from individualized electrode locations [184,190].

Extraneous movements, which are common [191], can also create EEG artifacts [e.g. [182],] and may make it difficult to focus on the BCI display. Further, gaze or visual impairments including ptosis (drooping) of the eye lid, nystagmus, and cerebral visual impairment (CVI) can lead to difficulty interpreting visual stimuli [192] for an SSVEP or P300 BCI device or visual feedback for an SMR BCI. Thus, special care is needed to understand how well the user can interpret visually presented information.

Indeed, user-centered design is important throughout BCI design and user training. Acclimation regimes may be needed with step-by-step introduction of individual BCI concepts. Family interactions, cooperation, and competition can increase motivation and engagement, which are essential for learning, but not a guarantor of good performance [193]. These factors are crucial as people with CP may have a long history of unsuccessful attempts to operate technology. Thus, the ideal BCI would be calibrated without the user following instructions, have intuitive operation, and be inherently engaging. In addition, systems should build on familiar concepts, such as row-column scanning, to simplify the transition from calibration to end-use [183].

Ultimately, we need improved understanding of the effect of CP on EEG, user-centered design to match the BCI to the interest and needs of individual users, and user-tailored training paradigms. Finally, it is vital to recognize that for children with congenital disabilities, technology use and even communication itself, are skills that must be taught.

3.3. From speech decoding to speech neuroprostheses

Organizer: Christian Herff (Maastricht University) and Sergey Stavisky (University of California, Davis)

Additional Presenters: Jon Brumberg (Kansas University), Phil Kennedy (Neural Signals Inc.), Miguel Angrick (University of Bremen), Julia Berezutskaya (Radboud University), Qinwan Rabbani (Johns Hopkins University)

Despite impressive recent results in decoding speech from neural recordings, there remain many challenges to achieving a real-time, large-vocabulary BCI for restoring lost speech. In this workshop, five of these challenges, and potential solutions, were discussed.

First, existing speech decoding demonstrations have not yet achieved consistently intelligible outputs. Multiple groups presented new decoding architectures, including recurrent neural networks and GANs. Workshop participants agreed that these modern machine learning approaches should benefit from additional data in future studies, and noted that all of the work presented used less than 20 minutes of neural recordings. Further, their performance did not saturate with training data quantity subsampled within these limited datasets.

A second challenge is how to obtain highly informative neural correlates about speech intent. Previous research almost exclusively relied on ECoG signals, which are not regularly used for long-term measurement. However, high-quality speech decoding and synthesis can also be achieved using penetrating microarrays implanted in the dorsal motor cortex [194], even though that area is not typically associated with speech production [195]. These Utah arrays have been used for multiple-year recordings in a number of participants and achieved high performance in, e.g. online decoding of attempted handwriting in people with tetraplegia [196] or speech perception decoding [197]. Alternatively, stereotactic EEG, which is very similar to Deep Brain Stimulation electrodes [198] that routinely remain implanted for decades, was proposed for high-quality speech synthesis. The neurotrophic electrode, an entirely different type of electrode with good long-term potential [199], was also proposed for speech neuroprosthesis [200].

Third, a functioning neuroprosthesis needs to generate or decode speech in or near real-time [45]. However, previous studies demonstrating speech synthesis [44,201] or speech recognition [202,203] from ECoG data have primarily (except for [204,205]) been done offline on previously recorded overt or whispered speech. Approaches that process and decode intracranial EEG in real-time will provide direct feedback to the

patient. This has been done using imagined speech processes [206], building on prior work such as [207]. Recent progress toward a low latency (250 ms) ECoG speech synthesis pipeline shows proof-of-concept open-loop results. A non-invasive EEG neuroprosthesis based on an artificial vocal tract model [207] provides auditory and visual feedback to the user and might therefore help train speech neuroprosthesis users and pilot online speech BCI methods.

Fourth, the field would benefit from better speech synthesis performance metrics. Recent works typically uses variants on measuring correlation between true and decoded audio (e.g. for spectral or pitch features), which are poor proxies for intelligibility. Workshop participants agreed that adopting subjective intelligibility metrics is important, but this may need to wait until decoding performance is good enough for these metrics to become relevant (or else they will suffer from floor effects).

Fifth, all presenters agreed that data sharing is key to accelerating progress. One recently shared large dataset of speech perception in fMRI, ECoG, and sEEG, along with the associated impressive reconstruction quality provides the public research community with a fully annotated dataset [208].

3.4. Brain-computer interfaces for the assessment of patients with disorders of consciousness

Organizer: Christoph Guger (g.tec Guger Technologies OG)

Additional Presenters: Damien Coyle, (Ulster University), Kyousuke Kamada, (Hokashin Group Megumino Hospital), Rossella Spataro, (University of Palermo), Jing Jin, (East China University), Steven Laureys, (Brain Centre & GIGA Consciousness, Coma Science Group, University and University Hospital of Liege, Belgium; International Disorders of Consciousness Institute, Hangzhou Normal University, China; CERVO Brain Research, U Laval)

Bedside evaluation to assess conscious awareness after coma requires inferences based on patients' motor responsiveness [209] with limited diagnostic precision and prognostic information, increasing the ethical difficulty of decisions on life-prolonging therapies. Technologies such as functional neuroimaging and BCIs provide objective tools for diagnostic, prognostic, and therapeutic purposes [210]. About two-thirds of patients clinically diagnosed with 'unresponsive wakefulness syndrome (UWS)' (or 'persistent vegetative state') may show residual brain activity in PET studies [211] and are hence actually in a minimally conscious state (MCS) with a better chance of recovery.

BCIs can help reduce the diagnostic and prognostic uncertainty of both acute and chronic disorders of consciousness [212,213]. BCI should first be used to establish a reliable and reproducible response to a simple command. Then, one can attempt functional communication with simple yes/no questions and eventually spelling or message creation [212,213]. The mindBEAGLE (g.tec medical engineering GmbH) uses auditory P300, vibro-tactile P300 and motor imagery paradigms for these steps and rehabilitation protocols. Paradigms include a quick (2–8 minute) system calibration or patient assessment. Other BCI systems have also been designed for this purpose, including using auditory sensorimotor rhythm feedback for those with visual impairments [214,215].

BCI assessment of DOC with locked-in and completely locked-in patients found 9 out of 12 patients could demonstrate command following by answering YES/NO questions [216]. Building on the pilot of 15 patients reported in [215], the workshop reported an update with 25 patients who each participated in 10, one-hour motor imagery BCI sessions. Of these, 5/9 UWS, 7/11 MCS, and 3/4 locked-in syndrome demonstrated significant capacity to modulate brain activity in stage I (assessment) and progressed to stage II/III (auditory feedback training and Q&A response). All participants in stage II/III responded significantly to YES/NO questions. Another study with unresponsive patients showed 3 out of 12 patients could successfully answer the YES/NO questions on some assessment days [217], showing that these patients have fluctuations in consciousness that can be detected by BCI systems.

BCIs can also help predict eventual recovery. Auditory P300 and vibro-tactile P300 provided a predictor of functional recovery for two patients with DOC. One patient did not show any auditory P300 or vibro-tactile P300 after three weeks and coma continued for more than 6 months. A second patient responded to auditory P300 and vibro-tactile P300 and after 6 months had recovered from coma and understood verbal commands. Such patients may benefit not only from BCI assessment, but also from BCI-based rehabilitation [218]. Longitudinal observation of 12 DOC patients showed that achieving mindBEAGLE classification accuracy of at least 50% predicts recovery of behavioral responsiveness (after six months) as measured by the coma-recovery scale revised (CRS-R) [219]. Moreover, 12 of 20 patients showed CRS-R score improvement after 10 sessions of a vibrotactile stimulation protocol [218].

BCI can also evaluate the effectiveness of other treatments for arousing DOC patients by analyzing EEG recorded during mental tasks before and after

intervention. BCI methods have been used to assess the effectiveness of spinal cord stimulation and deep brain stimulation surgeries in arousing vegetative patients. Auditory, vibro-tactile, or motor imagery-based BCI systems have been used to assess five unresponsive patients and three vegetative patients in this on-going study.

BCIs are being cross-validated against neuroimaging techniques such as PET and fMRI [220]. The current challenge is to integrate BCIs with our increasing scientific understanding of recovery from severe brain injury to optimize the trajectory of clinical care after coma and improve the quality-of-life in disorders of consciousness and locked-in syndrome [221].

3.5. The promise of BCI-driven functional recovery after stroke: leveraging current evidence to define next steps

Organizer: A Nicole Dusang (Brown University/ Providence VA Medical Center/Massachusetts General Hospital)

Additional Presenters: Murat Akcakaya (University of Pittsburgh); Febo Cincotti (Sapienza University); Cuntai Guan (Nanyang Technological University); Christoph Guger (g.tec medical engineering GmbH); Kyousuke Kamada (Asahikawa Medical University); David Lin (Massachusetts General Hospital/ Providence VA Medical Center); Donatella Mattia (Fondazione Santa Lucia IRCCS); José del R. Millán (University of Texas at Austin); Ander Ramos-Murguialday (University of Tübingen/TECNALIA Research and Innovation); Vivek Prabhakaran (University of Wisconsin-Madison); and George F. Wittenberg (Pittsburgh VA Healthcare System/ University of Pittsburgh)

Stroke is a leading cause of long-term disability worldwide, and 30–50% of stroke patients experience limited recovery. Rehabilitative EEG-BCIs are a promising neurotechnology for restoration of function after stroke. The hypothesis behind rehabilitative BCIs is that coupling neural activity with sensory feedback of limb movement induces cortical plasticity, improving functional recovery. This workshop featured 12 researchers developing rehabilitative EEG-BCIs for functional recovery from 10 institutions around the globe. Presenters were split into two panels to consider how to translate this technology from the lab to the clinic. Randomized controlled trials (RCTs) have demonstrated the benefit of Rehabilitative EEG-BCIs but employed diverse control methods, therapy doses, dosing intervals, and different types of neural dynamics and sensory feedback.

Panel 1 discussed optimal EEG-BCI support for stroke rehabilitation. Spatial neglect is an often overlooked deficit in stroke patients though it can significantly impact a patient's response to therapeutic intervention [222]. Technology is needed to objectively map neglect, quantify changes during recovery, and provide a rehabilitation platform to target spatial neglect. Although BCI addresses a gap in standard neurorehabilitation medicine [223], it still lacks an American Heart Association (AHA) class and evidence rating. BCIs empirically measure the signals of the damaged cortex and patients' functional disability during recovery. Rehabilitative EEG-BCIs restore the neural activity-functional output connection, supporting the retraining of neural activity. This is demonstrated by an RCT evaluating an EEG-BCI intervention for distal upper extremity function in a chronic stroke population [224]. Results showed 64% of participants made significant gains in both primary and secondary outcome measures.

Panel 2 reflected on stakeholders' needs for translating this promising technology to a clinical environment. Although RCTs have demonstrated the therapeutic efficacy of rehabilitative EEG-BCIs, commercialization requires clear clinical and economic benefit and reliable function within the rigors and environment of long-term clinical use. BCI-FES systems must address both patients' and clinicians' needs [225]. Patients need an effective and engaging rehabilitation platform, while clinicians require a plug-n-play system with remote technical assistance and joint analysis. Unanswered questions remain along the spectrum of basic research to patient care [226]. The field has yet to determine the optimal neural modalities or features for rehabilitative EEG-BCIs, resulting in significant feature extraction variability in current EEG-BCI platforms. Additionally, past and current RCTs employed diverse outcome measures since no measure is clearly best for capturing recovery. Further, stroke is itself a heterogeneous condition and much remains unknown about the relationship between the type and location of damage and resulting deficits. The RecoveriX system (Guger Technologies), a certified medical product, analyzes motor imagery to trigger FES for upper and/or lower limbs. RecoveriX has shown effectiveness for spasticity reduction and movement restoration in upper and lower limbs [227,228].

Convincing clinicians, patients, and payers that Rehabilitative BCIs are a worthy technology for investment was felt to require a large, multi-site, randomized control trial study, incorporating methods to minimize, or scientifically account for, heterogeneity between

technology and control populations at various sites. Ideally, it will also address knowledge gaps such as long-term effects, dose-response curves, patient stratification, control features, and a comprehensive outcome evaluation.

3.6. Towards the decoding of neural information for motor control: present and future approaches

Organizer: Gernot Müller-Putz (Graz University of Technology)

Additional Presenters: Andrea I. Sburlea (Graz University of Technology), Valeria Mondini (Graz University of Technology), Damien Coyle (Ulster University), Cuntai Guan (NTU Singapore), Tonio Ball (University of Freiburg)

For people with a cervical spinal cord injury (SCI) from trauma or disease, upper extremity function is often reduced or lost, resulting in dependency on a caregiver or family member for most daily activities. BCI researchers have for decades worked to derive motor commands directly from brain activity to bypass the interrupted spinal cord pathways and establish direct control of a neuroprosthetics device [229] or robotic arm/exoskeleton [230]. Implantable BCI approaches have produced many advances [231,232], however, in recent years, non-invasive approaches have moved beyond proof of concepts [233–235] and made major steps towards full arm control. This workshop focused on state-of-the-art approaches to non-invasive neural control of movement.

Non-invasive detection of multiple types of hand movements have been reported, including for people with cervical SCI [236,237]. Analysis of movement-related cortical potentials (MRCP) can detect and decode single hand movements [238] or movement attempts (e.g. hand open vs. hand close) or even different grasps (e.g. palmar vs. lateral grasp) [239,240].

Understanding the neural and behavioral mechanisms involved in grasping is important for successful decoding. Investigations included the relationship between the broad-band EEG representation of observing and executing a large variety of hand-object interactions and the muscle and kinematic representations associated with the grasping execution [241]. Object properties and grasp types can be decoded during the planning and execution of the movement. Properties of the objects could be decoded even during the observation stage, while the grasp type could be accurately decoded even during the object release stage [242].

While the decoding of arm/hand trajectories has mainly been shown in intracortical recordings, major steps in the non-invasive field have been demonstrated.

Closed-loop continuous decoding of executed [243,244] but also attempted arm movement [245] has been done from low-frequency EEG. Movement parameters like position and velocity, necessary for decoding [246,247] were presented. In particular, the contribution of non-directional movement-parameters (distance and speed) has been highlighted [248–250]. Also, the first evidence for online decoding of attempted continuous movement has been reported [245]. Eye movement artifacts present a special challenge for all non-invasive decoding studies. Participants must be permitted to use their gaze to follow the feedback, electroc-oculogram (EOG) signals must therefore be removed from the EEG online [251].

In addition to decode of low-frequency EEG components, decoding of executed and imagined 3D reaching tasks have involved delta frequencies, but also alpha, low and high beta frequencies [252,253]. These studies include decoding of 3D lower limb movements that could be important for gait rehabilitation [254].

In the area of motor imagery and stroke rehabilitation, deep learning methods and convolutional neural networks (CNN) have been used for participant specific [255,256], participant-independent [257], and adaptive classifiers [258]. CNNs have also been used in assistive robot control with online adaptive motor classification [259].

Beyond the pure application of CNNs for decoding [98], the internal data representation and the effects of hidden unit activations provide possible insights into what the units of such networks learn and the possible hierarchical organization of spectral features [260]. These first insights may open a new way of understanding brain processes.

3.7. Biomimetic approaches to restore somatosensation

Organizer: Robert Gaunt (University of Pittsburgh)

Additional Presenters: Sliman Bensmaia (University of Chicago), Karthik Kumaravelu (Duke University), Alberto Mazzoni (Scuola Superiore Sant’Anna), Emily Graczyk (Case Western Reserve University), Luke Bashford (California Institute of Technology), Chris Hughes (University of Pittsburgh)

Rapid advances in BCI capabilities to decode and restore upper limb motor functions [261] often ignore the accompanying sensory losses. Strategies to restore somatosensation include intracortical microstimulation [262,263], cortical epidural stimulation [264–266], peripheral nerve stimulation [267–269] and spinal cord stimulation [270]. Regardless of approach, it is difficult to select stimulus parameters that improve the quality of conscious percepts and maximize functional

capabilities. This workshop explored the idea of using biomimicry as a framework to create stimulus trains. Biomimetic stimulation leverages knowledge of intact somatosensory neurophysiology with the intuition that stimulation parameters that evoke patterns of neural activity that match normal patterns will improve perception and function.

Decades of work characterizing skin mechanoreceptor responses in the hand during object manipulation [271] were integrated into TouchSim to accurately simulate primary afferent responses to a mechanical input [272]. The simulated population-level activity resembles the spatiotemporal dynamics of somatosensory neurons in the cortex during the same mechanical stimuli [273], with large transient signals at contact onset and offset [271,274]. However, simply replacing recorded or simulated spikes with stimulation pulses does not replicate the sensation. Additional computations are required to address anatomical complexities and electrical stimulation biophysics. A simulation platform using genetic algorithms and finite element models of the cortex, populated with realistic neurons, was developed to address these complexities [275]. Critically, the stimulus trains created through simulation more faithfully represented the desired cortical activity than stimulus trains designed using standard methods.

The utility of this computational tool and the principles of biomimicry were tested in peripheral nerve stimulation experiments in amputees. As a baseline, linear stimulation encoding schemes that did not capture important features of natural neural coding were effectively used by participants [267]. Similarly, event-based stimulation encoding that mimicked the natural onset-offset dynamics of primary afferents was also effective [276]. However, in a direct comparison, TouchSim was used to create multiple stimulation trains that were increasingly biomimetic. The most natural sensations were obtained with the stimulus trains that maximized biomimicry [277]. In other experiments, early work suggested that a particular biomimetic train could improve naturalness [269]. Upon repetition, and despite considerable effort to combine modeled fascicle recruitment with biomimetic and non-biomimetic stimulation trains, just two of five participants reported more natural sensation using biomimetic trains, highlighting the limitations of single-subject studies of perception.

Two different aspects of biomimicry were explored in human intracortical BCIs. Motor imagery and actual movement evoke similar brain activity. To explore this concept for somatosensation, neural activity patterns were recorded in somatosensory cortex and the supramarginal gyrus during imagined sensations [278]. Different imagined sensations were encoded stably in

the somatosensory cortex, suggesting that imagined sensation could guide stimulus train design, even in people left insensate from their injury. Finally, in a direct test of biomimetic principles, intracortical stimulus trains using fixed amplitudes and frequencies were compared to trains with stimulation amplitudes modulated by cortical activity patterns recorded from non-human primates [274]. The participant frequently rated the biomimetic trains as more natural, especially when the overall intensity was matched.

In summary, biomimicry is a principled and likely fruitful approach to create stimulation trains to restore somatosensation. Simulation and modeling tools can help design these trains, which have outperformed less realistic trains in both the peripheral and central nervous systems. Nevertheless, considerable development is still necessary, and these results must be validated in larger numbers of participants.

4. Expanding BCI Usability and availability

4.1. Toward an international consensus on user characterization and BCI outcomes in settings of daily living

Organizers: Mariska Vansteensel (UMC Utrecht) and Nataliya Kosmyna (Massachusetts Institute of Technology)

Additional Presenters: Andrew Geronimo (Department of Neurosurgery, Penn State College of Medicine, Hershey, PA, USA), Katya Hill (AAC-BCI iNNOVATION LAB, University of Pittsburgh, Pittsburgh, PA, USA), Theresa Vaughan (National Center for Adaptive Neurotechnologies, Stratton VA Medical Center, Albany, NY, USA)

BCI research is growing fast, and implantable and non-invasive communication-BCIs are being introduced to people with significant motor disability for independent use in daily living situations e.g. [42,279–286], allowing end-users to participate in research and development experiments and provide critical input into iterative user-centered design [287]. Such studies are crucial for the development of usable communication-BCIs and for their eventual widespread implementation to resolve the communication problems of people with diseases such as amyotrophic lateral sclerosis. However, most studies include only limited numbers of participants. Since the target user population for communication-BCIs is relatively small [288], large studies may not actually be possible. For translation of communication-BCIs to practical use, it is therefore essential to compare results across studies and in this way learn about environmental and participant/user characteristics affecting BCI performance

e.g. [289–291]. and the different usability perspectives of users, caregivers, and other stakeholders. Such comparison will strongly benefit from standardized reporting about users/participants and their environment, and from the use of similar metrics to assess BCI performance and outcome [292]. This workshop was designed to initiate a consensus list of reporting recommendations, specifically directed at the use of communication-BCIs in the daily life settings of people with significant motor disability. After brief presentations to introduce the topics of discussion [196,293–302], workshop participants shared their experiences and built consensus in breakout rooms. Key outcomes of these discussions include:

- (1) **Standardization is hard.** Standardization is a hard and complex task. Part of this complexity comes from the different focus areas of experiments designed by different disciplines.
- (2) **Age group matters.** Adult and pediatric BCI users need different training procedures and different primary outcome measures. But researchers need as much comparison as possible.
- (3) **Meeting users' end goals is paramount.** For any system to be introduced in their environment, end-users should be strongly involved in BCI design, goal setting, and outcome measure selection. Even existing standard metrics for reporting BCI system performance must be adapted to the goals of the end-user.
- (4) **Needs of primary users and their caregiver(s) may be different.** A BCI has multiple types of end-users and researchers must report on how well a BCI meets the needs and goals of both primary and secondary (e.g. caregivers) users.
- (5) **Different tasks produce different outcomes.** BCI outcome measures should consider the importance of each task to be conducted with the BCI, as well as the desired and accomplished frequency of conducting each task.
- (6) **Fatigue strongly affects BCI performance.** Both cognitive and physical fatigue need to be assessed and reported on.
- (7) **Medication can affect brain signals.** The effect of medication should not be underestimated, but medication use is seldom reported in papers.

As our next steps, we plan to engage in the bigger discussion about standardization, to collect more input from BCI researchers, and to use all collected information for a formal publication on reporting recommendations related to user characterization and outcome measures for the use-case of communication-BCIs in settings of daily living.

4.2. On the need of good practices and standards for Benchmarking Brain–Machine Interfaces

Organizer: Ricardo Chavarriaga (Zurich University Applied Sciences, ZHAW Switzerland)

Additional Presenters: Paul Sajda (Columbia University, USA), José Contreras-Vidal (IUCRC BRAIN, University of Houston, USA), Luigi Bianchi (‘Tor Vergata’ University of Rome, Italy), Zach McKinney (Scuola Superiore Sant’Anna, Italy), Laura Y. Cabrera (The Pennsylvania State University, USA)

Translating Brain–Machine Interface (BMI) systems onto real applications requires accepted, well-defined criteria to assess their effectiveness, usability, and safety. Benchmarking, specification, and performance evaluation are perceived as main priorities for standardization in the field [292,303,304]. This workshop discussed translational challenges, and ethical issues of BMI systems, as well as existing initiatives to address them.

The Future Neural Therapeutics technology roadmap [305] analyzes closed-loop neurotechnologies aimed at treating movement disorders and neurological diseases. This document summarizes the state of the art and identifies key technological challenges required to successfully develop a new generation of these technologies, including computational power, robustness and safety, usability and appropriate regulatory frameworks. As BMIs approach commercial availability, attention must be paid to concerns generated by the possibility of repurposing, misusing, or maliciously using consumer-oriented neurotechnology. These concerns include overstated claims on their efficacy or the influence of neurotechnology in markets related to employment or cognitive enhancement [306–308]. Moreover, widespread use of consumer-oriented technology can lead to indiscriminate collection of neural data or user harm due to maladaptive processes triggered by neurostimulation devices.

The neuroethics subcommittee of the IEEE Brain Initiative focuses on the ethical and societal issues related to research and development of neurotechnologies. They developed the IEEE Neuroethics Framework (<https://brain.ieee.org/publications/ieee-neuroethics-framework/>), a collective effort to evaluate the ethical, legal, social, and cultural issues that arise with the deployment of neurotechnologies and provide explicit guidance on how to address them. The framework is organized as a matrix that covers existing and emerging neurotechnologies for both current and foreseen applications. This framework is conceived as a living document that will evolve with the technology. Participation in this effort is open to interested participants.

Despite the large number of BMI publications, it is seldom possible to evaluate, verify, or compare published results. Meta-analyses showed that a significant number of BCI publications lack necessary information [309,310]. However, two standardization activities are addressing this issue. The IEEE Standards Working Group P2794: *Reporting Standard for in vivo Neural Interface Research* (RSNIR) (<https://sagroups.ieee.org/2794/>) aims to improve the transparency, interpretability, and replicability of neural interface research by specifying a set of technological and methodological characteristics to be reported in scientific literature and technical documentation.

They recently published a set of preliminary requirements for implantable neural interfaces [311] and are seeking broad community input and participation to ensure the Standard reflects the needs of a more diverse range of neuroscience and neurotechnology stakeholders, including device regulators, funding officers, clinicians, and end users. Information on providing such input can be found through the working group website. Another standardization project, IEEE P2731: *Standard for a Unified Terminology for Brain–Computer Interfaces (BCI)* (<https://sagroups.ieee.org/2731/>) aims at developing a comprehensive BCI lexicography and a functional model of BCI systems [312–314]. It is also working on identifying the required information to be stored in BCI files to enable efficient sharing of data and tools among stakeholders [315]. These activities can contribute to the development of standard experimental and usage protocols, benchmarking procedures, and increased interoperability of neurotechnology systems.

Overall, this workshop highlighted the need to continuously evaluate the state-of-the-art and the implications of neurotechnologies. This requires multi-stakeholder, anticipatory processes for developing appropriate tools including ethical and technical guidelines, standards, and regulatory instruments that allow translation of neurotechnologies for both consumer and medical applications [316–318].

4.3. Lessons from successfully implanted neurotechnology

Organizer: Erik Aarnoutse (Brain Center, University Medical Center Utrecht)

Additional Presenters: Fabien Sauter-Starace (CEA, LETI, Cinatec, University of Grenoble); Leigh Hochberg (Brown University; Massachusetts General Hospital; Providence VA Medical Center), RI Aysegul Gunduz (J. Crayton Pruitt Family Department of Biomedical Engineering, University of Florida)

Over the last 16 years, various clinical trials of implantable neurotechnology in humans have demonstrated successful applications. This technology has enabled users to move arms [319,320], walk [108], and communicate [42] and has also alleviated disease symptoms [61]. Clinical trials require a great deal of effort but are an important and informative step along the route to wide availability of neurotechnology for users in need.

The route from design to clinical trial was illustrated by the Wimage implant to operate an exoskeleton [108]. First, the medical needs of people with quadriplegia were combined with the neurosurgical requirements: no transcutaneous connection, no limit to battery lifetime and limited invasiveness. This created design choices of wireless data transmission, inductive charging, and epidural ECoG electrodes. Technical requirements were a trade-off between wishes and constraints. Animal studies assessed signal stability [321]. Regulatory compliance to the EU Medical Device Regulation meant proving compliance to ISO standards for quality management and standards for mechanical, electrical, and thermal safety, biocompatibility, and software. The clinical trial with bilateral implants has enrolled two patients so far [108]. Training was progressive by adding more complexity in the adaptive machine learning algorithm, from brain switch to 3D + pronation/supination [322]. The signal proved to be stable over months. The exoskeleton was only used in the laboratory.

The 17 years of BCI research with penetrating multi-electrode arrays produced many lessons [319]. Participants are colleagues, but also customers. They request new features (user needs), which are added to the design [196,323]. The participants' motive is to advance science, they do not expect gain for themselves. However, the obligation of the field is to give users gain in daily life as soon as technology allows it [42,286]. Neuroethics is important here. Hardware advances ease the technical constraints making neural data ever easier to gather and use.

With the entry of industry in this field, the question of the role of academia becomes more important, where academia is better equipped to ask fundamental (hypothesis based) questions of neuroscience. Development is important but is not easy to publish. Mainly, academia investigates (hardware agnostic) decoding principles.

A good example of the input of academic expertise is seen in the use of cortical ECoG recordings as part of essential tremor DBS therapy [61]. This cross-field input produced knowledge on biomarkers both for fundamental questions and treatment efficacy. Here, user needs for individualized therapy, reduction of side

effects [324], and increased battery life were addressed. The research triggered a new hardware design that reduced stimulation artifacts.

Therefore, academia provides design input (user needs, technical requirements, decoding principles) for future neurotechnology for home use. Academia seeks to create knowledge, optimize designs, and provide a foundation of information that can support translation of BCI to commercial availability. We have also identified barriers that must be overcome for home use (wireless link, power constraints, limits on the number of electrodes, portability, larger scale manufacturing). Overcoming these barriers requires more time and money than academia has, but the generation of this knowledge by academic reduces the risk for industry and thus advances the likelihood that BCI will become widely, commercially available.

4.4. Next steps for practically useful BCI ethics

Organizer: Brendan Allison (UC San Diego)

Additional Presenters: Pim Haselager (Radboud University Nijmegen), Dr. Sonja Kleih-Dahms, (University of Würzburg), Donatella Mattia (Fondazione Santa Lucia, IRCCS)

This workshop was designed not for review or abstract academic discourse, but to develop practical next steps for BCI-related ethical issues. The organizers briefly presented examples of these issues [325–329] to promote discussion.

A public database of ethical use cases was proposed to raise awareness with an associated forum where people could share their perspectives on each case. The ethical use cases could also help professors and others who want to teach BCI ethics. Further discussion and development of ethical use cases would benefit from an ongoing collaborative effort, perhaps via online seminars, to develop a framework, assign people to develop different use cases, and create an online database. These efforts might be hosted by the BCI Society.

An immediate ethical concern is that research study participants do not usually keep the devices used in the study. Thus, people with disabilities may regain the ability to communicate or control a device with an experimental BCI, but then lose that ability when their study participation ends. Workshop contributors agreed that this is a serious and currently unresolved problem. Most funding sources do not support leaving devices with patients, nor providing ongoing technical support. However, several researchers include such considerations in their research plans. Possible next steps include

raising awareness of this problem (such as through an online forum, survey, paper, or approaching journalists) and further engagement of funding organizations.

The rise of 'Big BCI' through the recent initiation of BCI projects by high-profile companies creates its own set of ethical concerns. Workshop participants desired collaboration between the huge companies working on BCIs and the existing BCI community on efforts such as an online workshop or paper. This step was hoped to foster joint work on proposed ethical guidelines and regulatory issues.

Another concern comes from the many online articles and videos with misinformation about BCIs from different groups, including some manufacturers, neuro-feedback practitioners, enthusiasts, and conspiracy theorists. Of course, such misinformation will continue indefinitely to some extent, but might be reduced through next steps such as publicly commenting on inaccuracies and producing and promoting high-quality information about BCIs. Indeed, some for-profit and non-profit entities do provide good BCI content. The ongoing increase in online BCI-related classes, conferences, workshops, competitions, and other activities has led to ample recorded material from reputable organizers and speakers that is usually available for free.

Many participants had seen online postings from, and/or been directly contacted by, people who believe that they are being involuntarily mind-controlled by a BCI or a similar device. A few participants reported trying to direct such persons to appropriate mental health professionals, but without apparent success. Next steps at this time are not obvious aside from a possible paper or position statement with suggested guidelines, developed with mental health experts.

The workshop focused on specific, actionable next steps to raise awareness of ethical issues in BCI and further engage relevant groups through workshops, papers, online discussions and a database of use cases and surveys [330–332].

4.5. Brain–Computer Interfaces for human enhancement

Organizer: Davide Valeriani (Neurable Inc.)

Additional Presenters: Riccardo Poli (University of Essex), Maryam Shanechi (University of Southern California), Hasan Ayaz (Drexel University), Nataliya Kosmyna (MIT Media Lab), Yannick Roy (NeuroTechX), Marcello Ienca (ETH Zurich)

This workshop highlighted recent advances in BCI technologies that go beyond clinical applications and instead focus on augmenting human capabilities. The workshop brought together neuroscientists, engineers,

neuro-ethicists, entrepreneurs, and researchers at the cutting-edge of BCI development for human augmentation. Discussion focused on current trends and future prospects, as well as the critical role played by international communities such as NeuroTechX in educating and stimulating interest in BCI and neurotechnologies.

BCIs for cognitive human augmentation are intended to improve the process of acquiring knowledge and communicating with other individuals [333]. Passive BCIs can enhance individual decision-making in target detection by recognizing event-related potentials [334] or aggregating brain activity from multiple people [335]. Collaborative BCIs can also decode decision confidence from brain activity and use it to weigh individual opinions, leading to significant improvements in group performance in a variety of tasks [336–338]. These BCIs can also facilitate human-machine teaming in face recognition [339].

Combining brain recording (e.g. EEG, fNIRS) and stimulation (e.g. tDCS, TMS) improves processing speed [340] and spatial working memory [341], and introduces novel communication forms, such as brain-to-brain communication [342]. Moreover, it enables the development of BCIs capable of regulating abnormal mental states, with direct applications in the treatment of mental disorders [343,344].

BCIs and other wearables support studying the brain in complex environments and diverse domains, a research field called neuroergonomics [345]. Advances in recording technologies, such as EEG and fNIRS, enable study in operational and realistic settings to monitor cognitive function, improve human-to-human communication, and enhance human-machine interaction [346]. Moreover, the integration of brain recordings with other physiological signals can provide biofeedback to users through audio, light, or haptic inputs, promoting performance, attention, and overall well-being [347]. These hybrid, multimodal BCIs will also help increase the reliability, accuracy, and commercial potential of non-invasive BCIs, which can be limited by the low signal-to-noise ratio of non-invasive neural recordings. Yet to implement multimodal BCIs we need to identify relationships between modalities and develop new techniques to integrate neural recordings at different scales.

While neuroscience and neuro-engineering have shown that it is technically possible to develop BCIs that augment human capabilities in a variety of domains, neuro-ethicists are working to identify which applications are morally desirable [317]. Two main ethical principles should guide the development of BCIs for human augmentation: (1) cognitive liberty, which

protects the rights of individuals to make free and competent decisions on using such devices, and (2) fair and equitable access to enhancement, which ensures they are available to everyone, regardless of race, gender, or socioeconomic status. As with all biomedical devices, safety, and data privacy are key pillars to make these devices ethically acceptable.

Overall, the workshop showcased the tremendous advantages of expanding BCIs from assistive devices to technologies for human enhancement, with a variety of potential applications. The most promising approaches seem to be the fusion of different physiological signals and integration with artificial intelligence, with a continuous awareness of the ethical challenges of enhancement applications.

4.6. Brain–Computer Interfaces for outside the lab: neuroergonomics for human-computer interaction, education, and sport

Organizers: Antonia Thelen (eemagine Medical Imaging Solutions GmbH, Berlin, Germany)

Additional Presenters: Fabien Lotte, (Inria Bordeaux Sud-Ouest); Camille Jeunet (CNRS, Bordeaux Neurocampus); Frédéric Dehais (ISAE-SUPAERO, Toulouse); Patrique Fiedler (TU Ilmenau, Ilmenau); Martijn Schreuder (ANT-Neuro, Enschede)

Traditionally, BCI research has been bound to the investigation of perceptual, cognitive and motor processes within stationary, hardware-intensive laboratory setups. While these studies provide intriguing real-time insights into such processes, the translation of these findings into real-world brain interactions is limited. The emergence of lightweight, high-density EEG solutions has permitted the extension of BCI applications into mobile setups within real-world situations. Use of high-density EEG enables the simultaneous utilization of different sensor configurations, providing greater adaptability with a single hardware setup.

This workshop focused on the efforts undertaken toward the instrumentalization of EEG and specifically BCI techniques within the field of neuroergonomics. The panel comprised experts who strove to provide methodological strategies to facilitate the transition of BCI applications into real-world and/or every-day settings. First, advances and current limitations of existing solutions were discussed. Second, an outlook upon possible new technological and methodological innovations was presented which could provide new avenues of interacting with the world by implementing systems with an explicit awareness of the concepts of embodied cognition. Embodied cognition, as described in [348], acknowledges that physical elements of the world are

often integrated seamlessly into our cognitive processes in a way not easily captured by static diagrams with separate boxes for sensory inputs and physical outputs. Instead, cognition happens in conjunction and in parallel with the sensorimotor loops that provide interactions with the world. Various neuroergonomics applications of BCI use outside the lab were also discussed, including evaluating 3D User Interfaces [349], Sport Science [350,351] and Aviation [352].

Specifically, the robustness of signal processing methods used by BCI classifiers was discussed. How to apply such algorithms reliably across a large variety of application fields and how to make them cope with inter- and intra-individual variability is still a topic under investigation [353]. The contribution of state-of-the-art, lightweight, dry sensors resulting in varying signal-to-noise ratios and their impact upon such signal processing algorithms was highlighted [354,355]. Moreover, the tradeoff between laboratory-based and real-world applications was discussed with regard to sensor application within these fundamentally different environments [351,356]. Finally, discussion focused on difficulties encountered when translating BCI-based interventions across different demographics, specifically differences in cognitive states and/or perceptual processes that were investigated within a research context or focused on clinical/therapeutic interventions.

Taken together, the workshop provided an overview of current advances made within the field of neuroergonomics.

4.7. Brain–Computer Interfaces for art, entertainment, and domestic applications

Organizer: Anton Nijholt (University of Twente)

Additional Presenters: Christoph Guger (g.tec medical engineering GmbH); Elisabeth Hildt (Illinois Institute of Technology); Erika Mondria (University of Art and Design); Ellen Pearlman (Massachusetts Institute of Technology); Stephanie Scott (Colorado State University); Aleksander Valjamae (Tallinn University)

BCI technology enables neurophysiological data from an individual user's affective and mental state to be used for online adaption of system and interaction methods [357]. Artistic, domestic, or entertainment use of such information shift the focus from efficiency to the importance of affect in social and playful interactions such as in family, community, playful, and artistically challenging situations. This workshop addressed the use of BCI for artistic, entertainment, educational, and health applications.

BCI has been used for many artistic applications [358–360]. In general, artistic projects reduce inhibitions and encourage people to engage with unfamiliar technologies such as BCI. Synergies of design, art, and research have shown interesting results, which may also enrich clinical settings.

BCIs have been used for creative arts therapy [361,362] as part of a conceptual framework bringing together several disciplines for researching the expansion of treatment modalities in the intersection of art, technology, and therapeutics. A recent insight is that a post-phenomenological approach toward human-technology interaction and technological artifacts in general will be useful when applied to BCI for therapy, art, and creative expression. In this approach, user-specific needs for enabling self-expression are integrated in a transdisciplinary design perspective on meaningful and self-expressive communication exploring brain activity underlying artistic creation and using neuro-feedback research [363].

The BR41N.IO BCI Hackathon series, now in its 5th year [364,365], provides opportunities for team-based development of new BCI applications within 24 hours. During the 2021 BCI & Neurotechnology Spring School, 321 developers, artists, programmers, and hackers participated in 38 teams and created many interesting and cutting-edge new applications or improved the signal processing of BCI data sets.

In neurotheatre and neurocinema research [366,367], new media art and neurotechnologies allow for co-creation between actors, director, and audience to shape a performance by emotional experiences using BCI and other sensors and multisensory actuators. From a research perspective, neurotheatre can be seen as a novel integrative research environment for prototyping and exploring new social neuroscience paradigms, like collective decision-making or shared affective experiences. From a societal perspective, the fusion of science, technology, and arts allows for so-called design fiction, a design practice aiming at exploring and criticizing possible futures by creating speculative, and often provocative, scenarios narrated through designed artifacts.

Affective brain-computer music [368,369] Interface applications use affective BCIs for music-making and music listening. Given recent developments in direct-to-consumer devices (wearable BCIs, headphone sensors) and music streaming services these BCI applications aim at influencing the user's affective state (mood enhancement) by individualized music choices. Exaggerated claims about capabilities, increasing dependency on technology and limiting one's own capabilities, and

privacy issues arising from long-term monitoring of a user's affective state are pitfalls related to a potential future, relatively widespread use of EEG-based affective brain-computer music interfaces in entertainment contexts [370].

A brain opera called 'Noor' provides an example that combines these concepts through the use of artificial intelligence (AI). In 'Noor', biometric variables, including BCI are integrated with natural language processing and machine learning. In the near-future, such integrated systems will be tasked with more responsibilities relating to many aspects of human congress, often with confusing legal oversight and minimal accountability, potentially leading to scenarios enforcing dystopic digital societies of control [371–373].

The workshop discussions revealed consensus about the benefit of the joint effort of art and science research for BCI research in general and the acceptance of BCI for the general public.

5. Conclusion

Together, these workshops provide foundational information, explore diverse applications for different populations, and further develop big picture ideas for new frontiers of BCI use. Many of these ideas will be further developed in the workshops of the planned in-person Ninth International Brain-Computer Interface Meeting, currently scheduled for the summer of 2023 in the Sonian Forest, Brussels, Belgium.

Acknowledgments

Overall Acknowledgements

The authors thank the National Institute on Deafness and other Communication Disorders (NIDCD) and the National Institute of Neurological Disorders and Stroke (NINDS) in the National Institutes of Health (NIH) of the United States, the National Science Foundation (NSF), and the Wellcome Foundation for support assisting student participation in the BCI Meeting. We also thank the Research Foundation Flanders for support of the Meeting. The opinions expressed are those of the authors and do not reflect the views of any funding agency that may have supported work presented at the BCI Meeting or in the individual workshops.

The organizers of the workshop thank their presenters and participants for their presentations and thoughtful discussion. And, of course, we thank the many and varied funding sources that supported the research presented in the workshops. The workshop organizers also thank the members of the Program Committee for the Eighth International Brain-Computer Interface Meeting: José del R. Millán, Chuck Anderson, Guy Cheron, Jennifer Collinger, Marc Van Hulle, Dean Krusienski, Steven Laureys, Marc Slutzky, and Mariska Vansteensel.

Individual Workshop Acknowledgements

The work presented in *Biomimetic approaches to restore somatosensation* was supported by the NIH (R01 NS095251, U01 NS098975, UH3NS107714, U01NS108922), the National Science Foundation (IOS 1150209), the Kimberley Clark Foundation, EU Grant FET 611687, the Swiss National Science Foundation National Competence Center in Research in Robotics, the Bertarelli Foundation, the T&C Chen Brain-Machine Interface Center, the USC Neurorestoration Center, DARPA (NC66001-15-C-4041, N66001-16-C4501), and the US Department of Veterans Affairs (C3819C).

The workshop *Brain-computer interfaces for the assessment of patients with disorders of consciousness* was supported by the European Union's Horizon 2020 research and innovation programme under the Marie Skłodowska-Curie grant agreement No 841116 (RS).

The workshop *Brain-Computer Interfaces for Human Enhancement* was supported by the US DOD Bilateral Academic Research Initiative program (W911NF1810434). The funders had no role in workshop organization, decision to publish, or preparation of the manuscript.

The workshop *From Speech Decoding to Speech* was supported by NSF (1608140/1902395/2011595) and BMBF (01GQ1602) and as part of the NSF/NIH/BMBF Collaborative Research in Computational Neuroscience Program (CRCNS).

The workshop *On the need of good practices and standards for Benchmarking Brain-Machine Interfaces* was supported by the IEEE Brain Initiative, the IEEE Standards Association Industry Connections Program, the Confederation of Laboratories for Artificial Intelligence in Europe (CLAIRE), and National Science Foundation Awards #1650536 I/UCRC for Building Reliable Advances and Innovation in Neurotechnology (IUCRC BRAIN Center) and PFI # 1827769.

The workshop *The design of effective BCIs for children* was supported by the National Institute of Biomedical Imaging and Bioengineering of the NIH (P41 EB018783), resources at the US Department of Veterans Affairs Stratton VA Medical Center, the New York State Spinal Cord Injury Board, the Alberta Children's Hospital Foundation, and the Alberta Children's Hospital Research Institute (ACHRI).

The workshop *Toward the decoding of neural information for motor control: present and future approaches* was partly supported by the European Research Council (ERC-CoG -2015 681231 'Feel Your Reach')

Disclosure statement

CG is the owner and CEO of g.tec medical engineering GmbH. DV is an employee of Neurale, Inc. GMP is on the board of directors of the BCI Society. JEH is a co-Editor-in-Chief of *Brain-Computer Interfaces*, on the board of directors of the BCI Society, and has a pending patent on a BCI application used in one of the referenced papers. MV is on the board of directors of the BCI Society. RG is a member of the scientific advisory board for Braingrade GmbH.

Funding

This work was supported by the Alberta Children's Hospital Research Institute [#]; Bundesministerium für Bildung und Forschung [01GQ1602]; Defense Advanced Research Projects Agency [NC66001-15-C-4041, N66001-16-C4501]; European Research Council [ERC-CoG-2015 681231]; Fondation Bertarelli [#]; Horizon 2020 [841116]; IEEE Foundation [Brain Initiative and Standards Association Industry Connections Program]; National Institute of Biomedical Imaging and Bioengineering [P41 EB018783]; National Institutes of Health [UH3NS107714]; National Institutes of Health [U01NS108922]; National Institutes of Health [R01 NS095251]; National Institutes of Health [U01 NS098975]; National Science Foundation [IOS 1150209]; National Science Foundation [1650536 and PFI # 1827769]; National Science Foundation [1608140/1902395/2011595]; U.S. Department of Defense [W911NF1810434]; New York State Spinal Cord Injury Board [#]; Swiss National Science Foundation [#]; EU Grant FET [611687]; Alberta Children's Hospital Foundation [#]; Confederation of Laboratories for Artificial Intelligence in Europe [#]; Kimberley Clark Foundation [#]; National Competence Center in Research in Robotics [#]; u.s. department of veterans affairs [C3819C]; u.s. department of veterans affairs [#].

ORCID

Jane E. Huggins  <http://orcid.org/0000-0001-8709-4350>
 Dean Krusienski  <http://orcid.org/0000-0002-4668-5784>
 Mariska J. Vansteensel  <http://orcid.org/0000-0002-9252-5116>
 Davide Valeriani  <http://orcid.org/0000-0001-9866-0063>
 Antonia Thelen  <http://orcid.org/0000-0002-3989-5388>
 Sergey Stavisky  <http://orcid.org/0000-0002-5238-0573>
 Anton Nijholt  <http://orcid.org/0000-0002-5669-9290>
 Gernot Müller-Putz  <http://orcid.org/0000-0002-0087-3720>
 Nataliya Kosmyna  <http://orcid.org/0000-0003-1272-0470>
 Louis Korczowski  <http://orcid.org/0000-0002-7982-5739>
 Christoph Kapeller  <http://orcid.org/0000-0002-7330-895X>
 Christian Herff  <http://orcid.org/0000-0002-5610-2618>
 Sebastian Halder  <http://orcid.org/0000-0003-1017-3696>
 Christoph Guger  <http://orcid.org/0000-0001-6468-8500>
 Moritz Grosse-Wentrup  <http://orcid.org/0000-0001-9787-2291>
 Robert Gaunt  <http://orcid.org/0000-0001-6202-5818>
 Aliceson Nicole Dusang  <http://orcid.org/0000-0002-1480-1533>
 Ricardo Chavarriaga  <http://orcid.org/0000-0002-8879-2860>
 Brendan Allison  <http://orcid.org/0000-0001-7729-982X>
 Tetiana Aksenova  <http://orcid.org/0000-0003-4007-2343>
 Erik Aarnoutse  <http://orcid.org/0000-0001-7648-250X>

References

- [1] Wolpaw JR, Birbaumer N, Heetderks WJ, et al. Brain-computer interface technology: a review of the first international meeting. *IEEE Trans Rehabil Eng*. 2000;8(2):164–173. DOI:10.1109/TRE.2000.847807.

- [2] Vaughan TM, Heetderks WJ, Trejo LJ, et al. Guest editorial brain-computer interface technology: a review of the second international meeting. *IEEE Trans Neural Syst Rehabil Eng.* 2003;11(2):94–109. DOI:[10.1109/TNSRE.2003.814799](https://doi.org/10.1109/TNSRE.2003.814799).
- [3] Vaughan TM, Wolpaw JR. The third international meeting on brain-computer interface technology: making a difference. *IEEE Trans Neural Syst Rehabil Eng.* 2006;14(2):126–127.
- [4] Vaughan TM, Wolpaw JR. Special issue containing contributions from the fourth international brain-computer interface meeting. *J Neural Eng.* 2011;8(2):020201–2560/8/2/020201. Epub 2011 Mar 24. DOI:[10.1088/1741-2560/8/2/020201](https://doi.org/10.1088/1741-2560/8/2/020201).
- [5] Huggins JE, Guger C, Allison B, et al. Workshops of the fifth international brain-computer interface meeting: defining the future. *Brain-Comput Interface J.* 2014;1(1):27–49.
- [6] Huggins JE, Wolpaw JR. Papers from the fifth international brain-computer interface meeting. preface. *J Neural Eng.* 2014;11(3):030301–2560/11/3/030301. Epub 2014 May 19.
- [7] Daly JJ, Huggins JE. Brain-computer interface: current and emerging rehabilitation applications. *Arch Phys Med Rehabil.* 2015;96(3):S1–7.
- [8] Huggins JE, Guger C, Ziat M, et al. workshops of the sixth international brain-computer interface meeting: Brain-Computer Interfaces past, present, and future. *Brain Comput Interfaces (Abingdon).* 2017;4(1–2):3–36. DOI:[10.1080/2326263X.2016.1275488](https://doi.org/10.1080/2326263X.2016.1275488).
- [9] Huggins JE, Müller-Putz G, Wolpaw JR. The sixth international brain-computer interface meeting: advances in basic and clinical research. *Brain Comput Interfaces (Abingdon).* 2017;4(1–2):1–2.
- [10] Huggins JE, Guger C, Aarnoutse E, et al. Workshops of the seventh international brain-computer interface meeting: not getting lost in translation. *Brain Comput Interfaces (Abingdon).* 2019;6(3):71–101. DOI:[10.1080/2326263X.2019.1697163](https://doi.org/10.1080/2326263X.2019.1697163).
- [11] Huggins JE, Slutzky MW. Articles from the seventh international brain-computer interface meeting. *Brain-Comput Interfaces.* 2019;6(4):103–105.
- [12] Besio WG, Koka K, Aakula R, et al. Tri-polar concentric ring electrode development for laplacian electroencephalography. *IEEE Trans Biomed Eng.* 2006;53(5):926–933. DOI:[10.1109/TBME.2005.863887](https://doi.org/10.1109/TBME.2005.863887).
- [13] Koka K, Besio WG. Improvement of spatial selectivity and decrease of mutual information of tri-polar concentric ring electrodes. *J Neurosci Methods.* 2007;165(2):216–222.
- [14] Liu X, Makeyev O, Besio W. Improved spatial resolution of electroencephalogram using tripolar concentric ring electrode sensors. *J Sens.* 2020.
- [15] Besio WG, Martinez-Juarez IE, Makeyev O, et al. High-frequency oscillations recorded on the scalp of patients with epilepsy using tripolar concentric ring electrodes. *IEEE J Transl Eng Health Med.* 2014;2:2000111.
- [16] Toole C, Martinez-Juárez IE, Gaitanis JN, et al. Source localization of high frequency activity in tripolar electroencephalography of epilepsy patients. *Epilepsy Behav.* 2019;101:106519.
- [17] Besio W, Cao H, Zhou P. Application of tripolar concentric electrodes and pre-feature selection algorithm for brain-computer interface. *IEEE Trans Neural Syst Rehabil Eng.* 2008;16(2):191–194.
- [18] Boudria Y, Feltane A, Besio W. Significant improvement in one dimensional cursor control using laplacian electroencephalography over electroencephalography. *J Neural Eng.* 2014;11(3):35014.
- [19] Alzahrani SI, Anderson CW. A comparison of conventional and tri-polar eeg electrodes for decoding real and imaginary finger movements from one hand. *Int J Neural Syst.* 2021;31(9):2150036.
- [20] Besio W, Sharma V, Spaulding J. The effects of concentric ring electrode electrical stimulation on rat skin. *Ann Biomed Eng.* 2010;38(3):1111–1118.
- [21] Rogel-Salazar G, Luna-Munguía H, Stevens KE, et al. Transcranial focal electrical stimulation via tripolar concentric ring electrodes does not modify the short- and long-term memory formation in rats evaluated in the novel object recognition test. *Epilepsy Behav.* 2013;27(1):154–158. DOI:[10.1016/j.yebeh.2013.01.006](https://doi.org/10.1016/j.yebeh.2013.01.006).
- [22] Besio WG, Koka K, Cole AJ. Effects of noninvasive transcutaneous electrical stimulation via concentric ring electrodes on pilocarpine-induced status epilepticus in rats. *Epilepsia.* 2007;48(12):2273–2279.
- [23] Besio WG, Makeyev O, Medvedev A, et al. Effects of transcranial focal electrical stimulation via tripolar concentric ring electrodes on pentylenetetrazole-induced seizures in rats. *Epilepsy Res.* 2013;105(1–2):42–51. DOI:[10.1016/j.epilepsyres.2012.12.002](https://doi.org/10.1016/j.epilepsyres.2012.12.002).
- [24] Makeyev O, Liu X, Luna-Munguía H, et al. Toward a noninvasive automatic seizure control system in rats with transcranial focal stimulations via tripolar concentric ring electrodes. *IEEE Trans Neural Syst Rehabil Eng.* 2012;20(4):422–431. DOI:[10.1109/TNSRE.2012.2197865](https://doi.org/10.1109/TNSRE.2012.2197865).
- [25] Makeyev O, Luna-Munguía H, Rogel-Salazar G, et al. Noninvasive transcranial focal stimulation via tripolar concentric ring electrodes lessens behavioral seizure activity of recurrent pentylenetetrazole administrations in rats. *IEEE Trans Neural Syst Rehabil Eng.* 2013;21(3):383–390. DOI:[10.1109/TNSRE.2012.2198244](https://doi.org/10.1109/TNSRE.2012.2198244).
- [26] Besio WG, Liu X, Wang L, et al. Transcutaneous focal electrical stimulation via concentric ring electrodes reduces synchrony induced by pentylenetetrazole in beta and gamma bands in rats. *Int J Neural Syst.* 2011;21(2):139–149. DOI:[10.1142/S0129065711002729](https://doi.org/10.1142/S0129065711002729).
- [27] Valdes-Cruz A, Villasana-Salazar B, Williams B, et al. Transcranial focal electrical stimulation via concentric ring electrodes in freely moving cats: antiepileptogenic and postictal effects. *Exp Neurol.* 2019;320:113012.
- [28] Besio W, Cuellar-Herrera M, Luna-Munguía H, et al. Effects of transcranial focal electrical stimulation alone and associated with a sub-effective dose of diazepam on pilocarpine-induced status epilepticus and subsequent neuronal damage in rats. *Epilepsy Behav.* 2013;28(3):432–436. DOI:[10.1016/j.yebeh.2013.06.021](https://doi.org/10.1016/j.yebeh.2013.06.021).
- [29] Santana-Gomez CE, Alcántara-González D, Luna-Munguía H, et al. Transcranial focal electrical stimulation reduces the convulsive expression and amino acid

- release in the hippocampus during pilocarpine-induced status epilepticus in rats. *Epilepsy Behav.* **2015**;49:33–39.
- [30] Perez-Perez D, Castañeda-Cabral JL, Orozco-Suárez S, et al. Noninvasive transcranial focal stimulation affects the convulsive seizure-induced P-glycoprotein expression and function in rats. *Epilepsy Behav.* **2021**;115:107659.
- [31] Hansen N, Kahn AK, Zelle D, et al. Amplitudes of pain-related evoked potentials are useful to detect small fiber involvement in painful mixed fiber neuropathies in addition to quantitative sensory testing an electrophysiological study. *Front Neurol.* **2015**;6:244.
- [32] Oh KJ, Kim SH, Lee Y-H, et al. Pain-related evoked potential in healthy adults. *Ann Rehabil Med.* **2015**;39(1):108–115. DOI:10.5535/arm.2015.39.1.108.
- [33] Ozgul OS, Maier C, Enax-Krumova EK, et al. High test-retest-reliability of pain-related evoked potentials (PREP) in healthy subjects. *Neurosci Lett.* **2017**;647:110–116.
- [34] Papagianni A, Siedler G, Sommer C, et al. Capsaicin 8% patch reversibly reduces A-delta fiber evoked potential amplitudes. *Pain Rep.* **2018**;3(2):e644. DOI:10.1097/PR9.0000000000000644.
- [35] Uceyler N, Kahn A-K, Kramer D, et al. Impaired small fiber conduction in patients with fabry disease: a neurophysiological case-control study. *BMC Neurol.* **2013**;13(1):47. DOI:10.1186/1471-2377-13-47.
- [36] Uceyler N, Zeller D, Kahn A-K, et al. Small fibre pathology in patients with fibromyalgia syndrome. *Brain.* **2013**;136(6):1857–1867. DOI:10.1093/brain/awt053.
- [37] Katsarava Z, Ayzenberg I, Sack F, et al. A novel method of eliciting pain-related potentials by transcutaneous electrical stimulation. *Headache.* **2006**;46(10):1511–1517. DOI:10.1111/j.1526-4610.2006.00446.x.
- [38] Khatoun A, Asamoah B, Mc Laughlin M. Investigating the feasibility of epicranial cortical stimulation using concentric-ring electrodes: a novel minimally invasive neuromodulation method. *Front Neurosci.* **2019**;13:773.
- [39] Miller KJ, Hermes D, Staff NP. The current state of electrocorticography-based brain-computer interfaces. *Neurosurg Focus.* **2020**;49(1):E2.
- [40] Wang W, Collinger JL, Degenhart AD, et al. An electrocorticographic brain interface in an individual with tetraplegia. *PloS One.* **2013**;8(2):e55344. DOI:10.1371/journal.pone.0055344.
- [41] Brunner P, Ritaccio AL, Emrich JF, et al. Rapid communication with a “p300” matrix speller using electrocorticographic signals (ECoG). *Front Neurosci.* **2011**;5:5.
- [42] Vansteensel MJ, Pels EGM, Bleichner MG, et al. Fully implanted brain-computer interface in a locked-in patient with ALS. *N Engl J Med.* **2016**;375(21):2060–2066. DOI:10.1056/NEJMoa1608085.
- [43] Lotte F, Brumberg JS, Brunner P, et al. Electrocorticographic representations of segmental features in continuous speech. *Front Hum Neurosci.* **2015**;9:97.
- [44] Anumanchipalli GK, Chartier J, Chang EF. Speech synthesis from neural decoding of spoken sentences. *Nature.* **2019**;568(7753):493–498.
- [45] Rabbani Q, Milsap G, Crone NE. The potential for a speech brain–computer interface using chronic electrocorticography. *Neurotherapeutics.* **2019**;16(1):144–165.
- [46] Dijkstra K, Brunner P, Gunduz A, et al. Identifying the attended speaker using electrocorticographic (ECoG) signals. *Brain Comput Interfaces.* **2015**;2(4):161–173. (Abingdon, England). DOI:10.1080/2326263X.2015.1063363.
- [47] Miller KJ, Schalk G, Fetz EE, et al. Cortical activity during motor execution, motor imagery, and imagery-based online feedback. *Proc Natl Acad Sci U S A.* **2010**;107(9):4430–4435. DOI:10.1073/pnas.0913697107.
- [48] Miller KJ, Leuthardt EC, Schalk G, et al. Spectral changes in cortical surface potentials during motor movement. *J Neurosci.* **2007**;27(9):2424–2432. DOI:10.1523/JNEUROSCI.3886-06.2007.
- [49] Kapeller C, Ogawa H, Schalk G, et al. Real-time detection and discrimination of visual perception using electrocorticographic signals. *J Neural Eng.* **2018**;15(3):036001. DOI:10.1088/1741-2552/aaa9f6.
- [50] Miller KJ, Schalk G, Hermes D, et al. Spontaneous decoding of the timing and content of human object perception from cortical surface recordings reveals complementary information in the event-related potential and broadband spectral change. *PLoS Comput Biol.* **2016**;12(1):e1004660. DOI:10.1371/journal.pcbi.1004660.
- [51] Leuthardt EC, Moran DW, Mullen TR. Defining surgical terminology and risk for brain computer interface technologies. *Front Neurosci.* **2021**;15:599549.
- [52] Ogawa H, Kamada K, Kapeller C, et al. Clinical impact and implication of real-time oscillation analysis for language mapping. *World Neurosurg.* **2017**;97:123–131.
- [53] Kapeller C, Korostenskaja M, Prueckl R, et al. CortiQ-based real-time functional mapping for epilepsy surgery. *J Clin Neurophysiol.* **2015**;32(3):e12–22. DOI:10.1097/WNP.0000000000000131.
- [54] Ogawa H, Kamada K, Kapeller C, et al. Rapid and minimum invasive functional brain mapping by real-time visualization of high gamma activity during awake craniotomy. *World Neurosurg.* **2014**;82(5):912 e1–10. DOI:10.1016/j.wneu.2014.08.009.
- [55] Kohler F, Gkogkidis CA, Bentler C, et al. Closed-loop interaction with the cerebral cortex: a review of wireless implant technology. *Brain-Comput Interfaces.* **2017**;4(3):146–154. DOI:10.1080/2326263X.2017.1338011.
- [56] Kramer DR, Lamorie-Foote K, Barbaro M, et al. Utility and lower limits of frequency detection in surface electrode stimulation for somatosensory brain-computer interface in humans. *Neurosurg Focus.* **2020**;48(2):E2. DOI:10.3171/2019.11.FOCUS19696.
- [57] Schalk G, Kapeller C, Guger C, et al. Facephenes and rainbows: causal evidence for functional and anatomical specificity of face and color processing in the human brain. *Proc Natl Acad Sci U S A.* **2017**;114(46):12285–12290. DOI:10.1073/pnas.1713447114.
- [58] Brice J, McLellan L. Suppression of intention tremor by contingent deep-brain stimulation. *Lancet.* **1980**;315(8180):1221–1222.
- [59] Lozano AM, Lipsman N. Probing and regulating dysfunctional circuits using deep brain stimulation. *Neuron.* **2013**;77(3):406–424.
- [60] Molina R, Okun MS, Shute JB, et al. Report of a patient undergoing chronic responsive deep brain stimulation for tourette syndrome: proof of concept. *J Neurosurg.* **2018**;129(2):308–314. DOI:10.3171/2017.6.JNS17626.

- [61] Opri E, Cernera S, Molina R, et al. Chronic embedded cortico-thalamic closed-loop deep brain stimulation for the treatment of essential tremor. *Sci Transl Med*. 2020;12(572). DOI:10.1126/scitranslmed.aay7680.
- [62] Schalk G, McFarland DJ, Hinterberger T, et al. BCI2000: a general-purpose brain-computer interface (BCI) system. *IEEE Trans Biomed Eng*. 2004;51(6):1034–1043. DOI:10.1109/TBME.2004.827072.
- [63] Guger C, Schlögl A, Neuper C, et al. Rapid prototyping of an EEG-based brain-computer interface (BCI). *IEEE Trans Neural Syst Rehabil Eng*. 2001;9(1):49–58. DOI:10.1109/7333.918276.
- [64] Barachant A, Bonnet S, Congedo M, et al. Riemannian geometry applied to BCI classification. Springer; 2010.
- [65] Barachant A, Congedo M. A plug & play p300 bci using information geometry. *arXiv*. 2014;1409.0107.
- [66] Kalunga EK, Chevallier S, Barthélemy Q. Using riemannian geometry for ssvep-based brain computer interface. *arXiv:1501.03227 [cs, stat]*; 2015.
- [67] Li Y, Wong KM, Bruin HD. Electroencephalogram signals classification for sleepstate decision A riemannian geometry approach. *IET Signal Process*. 2012;6(4):288–299.
- [68] Simar C, Cebolla A-M, Chartier G, et al. Hyperscanning EEG and classification based on riemannian geometry for festive and violent mental state discrimination. *Front Neurosci*. 2020;14:1225.
- [69] Barachant A, Bonnet S, Congedo M, et al. Multiclass brain-computer interface classification by riemannian geometry. *IEEE Trans Biomed Eng*. 2012;59(4):920–928. DOI:10.1109/TBME.2011.2172210.
- [70] Kalunga EK, Chevallier S, Barthélemy Q. Transfer learning for ssvep-based bci using riemannian similarities between users. in 2018 26th European Signal Processing Conference (EUSIPCO); 2018:1685–1689.
- [71] Rodrigues PLC, Jutten C, Congedo M. Riemannian procrustes analysis: transfer learning for brain-computer interfaces. *IEEE Trans Biomed Eng*. 2019;66(8):2390–2401.
- [72] Korczowski L, Congedo M, Jutten C. Single-trial classification of multi-user P300-based brain-computer interface using riemannian geometry. in 2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC); 2015:1769–1772.
- [73] Barachant A. MEG decoding using Riemannian geometry and unsupervised classification. 2014. Available from: <http://citeseerx.ist.psu.edu/viewdoc/citations>.
- [74] Barachant A, Cycon R. Pushing the limits of BCI accuracy: winning solution of the grasp & lift EEG challenge. in 6th International Brain-Computer Interface Meeting, Pacific Grove, California, USA, 2016: 73.
- [75] Horev I, Yger F, Sugiyama M. Geometry-aware principal component analysis for symmetric positive definite matrices. *Mach Learn*. 2017;106(4):493–522.
- [76] Congedo M, Barachant A, Bhatia R. Riemannian geometry for EEG-based brain-computer interfaces; a primer and a review. *Brain-Comput Interfaces*. 2017;4(3):155–174.
- [77] Barachant A, Andreev A, Congedo M. The Riemannian potato: an automatic and adaptive artifact detection method for online experiments using Riemannian geometry. 2013: 19–20.
- [78] Barthélemy Q, Mayaud L, Ojeda D, et al. The riemannian potato field: a tool for online signal quality index of EEG. *IEEE Trans Neural Syst Rehabil Eng*. 2019;27(2):244–255. DOI:10.1109/TNSRE.2019.2893113.
- [79] Jayaram V, Barachant A. MOABB: trustworthy algorithm benchmarking for BCIs. *J Neural Eng*. 2018;15(6):066011.
- [80] Clisson P, Bertrand-Lalo R, Congedo M, et al. Timeflux: an open-source framework for the acquisition and near real-time processing of signal streams, in BCI 2019 - 8th International Brain-Computer Interface Conference, Graz, Austria; 2019.
- [81] Lotte F, Jeunet C, Chavarriaga R, et al. Turning negative into positives! Exploiting ‘negative’ results in brain-machine interface (BMI) research. *Brain-Comput Interfaces*. 2019;6(4):178–189. DOI:10.1080/2326263X.2019.1697143.
- [82] kaggle. 2018 Kaggle machine learning & data science survey. 2021 [cited 2021 July 10]. Available from: <https://kaggle.com/kaggle/kaggle-survey-2018>
- [83] Congedo M, Barachant A, Andreev A. A new generation of brain-computer interface based on riemannian geometry. *arXiv.org*; 2013. ArXiv:1310.8115 [Cs, Math].
- [84] Khazem S, Chevallier S, Barthélemy Q, et al. Minimizing subject-dependent calibration for bci with riemannian transfer learning. in 2021 10th International IEEE/EMBS Conference on Neural Engineering (NER); 2021.
- [85] PyRiemann/PyRiemann. Pyriemann. GitHub.2015 Available from: <https://github.com/pyRiemann/pyRiemann>
- [86] Lotte F, Bougrain L, Cichocki A, et al. A review of classification algorithms for EEG-based brain-computer interfaces: a 10 year update. *J Neural Eng*. 2018;15(3):031005. DOI:10.1088/1741-2552/aab2f2.
- [87] NeuroTechX. NeuroTechX/Moabb. 2017 [cited 2021]. Available from: <https://github.com/NeuroTechX/moabb>
- [88] Reback J, jbrockmendel W, McKinney J, et al. Pandas-Dev/Pandas: Pandas 1.3.0. v1.3.0. Zenodo. 2021 [cited 2021].
- [89] Pedregosa F, Varoquaux G, Gramfort A, et al. Scikit-learn: machine Learning in python. *J Mach Learn Res*. 2021;12(85):2825–2830.
- [90] Neuroscience, S.C.f.C. Sccn/Labstreaminglayer. 2018 [cited 2021]. Available from: <https://github.com/sccn/labstreaminglayer>
- [91] Group TH. The HDF5® library & file format. [cited 2021]. Available from: <https://www.hdfgroup.org/solutions/hdf5/>
- [92] Gramfort A, Luessi M, Larson E, et al. MEG and EEG data analysis with MNE-python. *Front Neurosci*. 2013;7:267.
- [93] Timeflux. Timeflux/Workshops. 2021.
- [94] Krizhevsky A, Sutskever I, Hinton GE. ImageNet classification with deep convolutional neural networks. *Commun ACM*. 2017;60(6):84–90.
- [95] Mnih V, Kavukcuoglu K, Silver D, et al. Human-level control through deep reinforcement learning. *Nature*. 2015;518(7540):529–533. DOI:10.1038/nature14236.
- [96] Silver D, Huang A, Maddison CJ, et al. Mastering the game of Go with deep neural networks and tree search. *Nature*. 2016;529(7587):484–489. DOI:10.1038/nature16961.

- [97] Senior AW, Evans R, Jumper J, et al. Improved protein structure prediction using potentials from deep learning. *Nature*. 2020;577(7792):706–710. DOI:10.1038/s41586-019-1923-7.
- [98] Schirrmeister RT, Springenberg JT, Fiederer LDJ, et al. Deep learning with convolutional neural networks for EEG decoding and visualization. *Hum Brain Mapp*. 2017;38(11):5391–5420. DOI:10.1002/hbm.23730.
- [99] Lawhern VJ, Solon AJ, Waytowich NR, et al. EEGNet: a compact convolutional neural network for EEG-based brain-computer interfaces. *J Neural Eng*. 2018;15(5):056013. DOI:10.1088/1741-2552/aace8c.
- [100] Muller KR, Anderson CW, Birch GE. Linear and non-linear methods for brain-computer interfaces. *IEEE Trans Neural Syst Rehabil Eng*. 2003;11(2):165–169.
- [101] Barachant A, Bonnet S, Congedo M, et al. Classification of covariance matrices using a riemannian-based kernel for BCI applications. *Neurocomputing*. 2013;112:172–178.
- [102] Jeunet C, Benaroch C, Cabestaing F, et al. A user-centred approach to unlock the potential of non-invasive bcis: an unprecedented international translational effort. in CHIST-ERA; 2020. virtual.
- [103] Deng J, Dong W, Socher R, et al. ImageNet: a large-scale hierarchical image database. in 2009 IEEE Conference on Computer Vision and Pattern Recognition; 2009.
- [104] Zander TO, Kothe C. Towards passive brain-computer interfaces: applying brain-computer interface technology to human-machine systems in general. *J Neural Eng*. 2011;8(2):025005–2560/8/2/025005. Epub 2011 Mar 24. DOI:10.1088/1741-2560/8/2/025005.
- [105] Jayaram V, Alamgir M, Altun Y, et al. Transfer learning in brain-computer interfaces. *IEEE Comput Intel Mag*. 2016;11(1):20–31. DOI:10.1109/MCI.2015.2501545.
- [106] Zanini P, Congedo M, Jutten C, et al. Transfer learning: a riemannian geometry framework with applications to brain-computer interfaces. *IEEE Trans Biomed Eng*. 2018;65(5):1107–1116. DOI:10.1109/TBME.2017.2742541.
- [107] Özdenizci O, Wang Y, Koike-Akino T, et al. Learning invariant representations from eeg via adversarial inference. *IEEE Access*. 2020;8:27074–27085.
- [108] Benabid AL, Costecalde T, Elisseyev A, et al. An exoskeleton controlled by an epidural wireless brain-machine interface in a tetraplegic patient: a proof-of-concept demonstration. *Lancet Neurol*. 2019;18(12):1112–1122. DOI:10.1016/S1474-4422(19)30321-7.
- [109] Elisseyev A, Auboiroux V, Costecalde T, et al. Recursive exponentially weighted n-way partial least squares regression with recursive-validation of hyper-parameters in brain-computer interface applications. *Sci Rep*. 2017;7(1):16281. DOI:10.1038/s41598-017-16579-9.
- [110] Schaeffer MC, Aksenova T. Switching markov decoders for asynchronous trajectory reconstruction from ECoG signals in monkeys for BCI applications. *J Physiol Paris*. 2016;110(4 Pt A):348–360.
- [111] Orsborn AL, Moorman H, Overduin S, et al. Closed-loop decoder adaptation shapes neural plasticity for skillful neuroprosthetic control. *Neuron*. 2014;82(6):1380–1393. DOI:10.1016/j.neuron.2014.04.048.
- [112] Madduri MM, Burden SA, Orsborn AL. A game-theoretic model for co-adaptive brain-machine interfaces. in 10th International IEEE/EMBS Conference on Neural Engineering (NER); 2021. p. 327–330.
- [113] Perdakis S, Leeb R, Williamson J, et al. Clinical evaluation of braintree, a motor imagery hybrid BCI speller. *J Neural Eng*. 2014;11(3):036003. DOI:10.1088/1741-2560/11/3/036003.
- [114] Perdakis S, Leeb R, Chavarriaga R, et al. Context-aware learning for generative models. *IEEE Trans Neural Netw Learn Syst*. 2021;32(8):3471–3483. DOI:10.1109/TNNLS.2020.3011671.
- [115] Perdakis S, Leeb R, Millan JD. Context-aware adaptive spelling in motor imagery BCI. *J Neural Eng*. 2016;13(3):036018.
- [116] Perdakis S, Millán JDR. Brain-machine interfaces: a tale of two learners. *IEEE Syst Man Cybernetics Magaz*. 2020;6(3):12–19.
- [117] Perdakis S, Tonin L, Saeedi S, et al. The cybathlon BCI race: successful longitudinal mutual learning with two tetraplegic users. *PLoS Biol*. 2018;16(5):e2003787. DOI:10.1371/journal.pbio.2003787.
- [118] Biasiucci A, Leeb R, Iturrate I, et al. Brain-actuated functional electrical stimulation elicits lasting arm motor recovery after stroke. *Nat Commun*. 2018;9(1):2421. DOI:10.1038/s41467-018-04673-z.
- [119] Amiri S, Fazel-Rezai R, Asadpour V. A review of hybrid brain-computer interface systems. *Adv Hum Comput Interact*. 2013;8:187024.
- [120] Armel KC, Ramachandran VS. Projecting sensations to external objects: evidence from skin conductance response. *Proc Biol Sci*. 2003;270(1523):1499–1506.
- [121] Critchley HD. Electrodermal responses: what happens in the brain. *Neuroscientist*. 2002;8(2):132–142.
- [122] Marini F, Tagliabue CF, Sposito AV, et al. Crossmodal representation of a functional robotic hand arises after extensive training in healthy participants. *Neuropsychologia*. 2014;53:178–186.
- [123] Shokur S, Gallo S, Moioli RC, et al. Assimilation of virtual legs and perception of floor texture by complete paraplegic patients receiving artificial tactile feedback. *Sci Rep*. 2016;6(1):32293. DOI:10.1038/srep32293.
- [124] Voss P, Thomas ME, Cisneros-Franco JM, et al. Dynamic brains and the changing rules of neuroplasticity: implications for learning and recovery. *Front Psychol*. 2017;8:1657.
- [125] de Vignemont F. Embodiment, ownership and disownership. *Conscious Cogn*. 2011;20(1):82–93.
- [126] Adama VS, Wu S-J, Nicolaou N, et al. Extendable hybrid approach to detect consciousness states in a CLIS patient using machine learning. in 10th EUROSIM Congress; Logroño, Spain; 2019.
- [127] Roussel P, Godais GL, Bocquet F, et al. Observation and assessment of acoustic contamination of electrophysiological brain signals during speech production and sound perception. *J Neural Eng*. 2020;17(5):056028. DOI:10.1088/1741-2552/abb25e.

- [128] Kubler A, Holz EM, Riccio A, et al. The user-centered design as novel perspective for evaluating the usability of BCI-controlled applications. *PloS One*. 2014;9(12): e112392.
- [129] Andrich R, Mathiassen N-E, Hoogerwerf E-J, et al. Service delivery systems for assistive technology in Europe: an AAATE/EASTIN position paper. *Technol Disabil*. 2013;25(3):127–146. DOI:10.3233/TAD-130381.
- [130] Liberati G, Pizzimenti A, Simone L, et al. Developing brain-computer interfaces from a user-centered perspective: assessing the needs of persons with amyotrophic lateral sclerosis, caregivers, and professionals. *Appl Ergon*. 2015;50:139–146.
- [131] Riccio A, Pichiorri F, Schettini F, et al. Interfacing brain with computer to improve communication and rehabilitation after brain damage. *Prog Brain Res*. 2016;228:357–387.
- [132] Riccio A, Simone L, Schettini F, et al. Attention and P300-based BCI performance in people with amyotrophic lateral sclerosis. *Front Hum Neurosci*. 2013;7:732.
- [133] Riccio A, Schettini F, Simone L, et al. On the relationship between attention processing and p300-based brain computer interface control in amyotrophic lateral sclerosis. *Front Hum Neurosci*. 2018;12:165.
- [134] Schreuder M, Riccio A, Riseti M, et al. User-centered design in brain-computer interfaces-a case study. *Artif Intell Med*. 2013;59(2):71–80. DOI:10.1016/j.artmed.2013.07.005.
- [135] Schettini F, Riccio A, Simone L, et al. Assistive device with conventional, alternative, and brain-computer interface inputs to enhance interaction with the environment for people with amyotrophic lateral sclerosis: a feasibility and usability study. *Arch Phys Med Rehabil*. 2015;96(3):S46–53. DOI:10.1016/j.apmr.2014.05.027.
- [136] Riccio A, Holz EM, Aricò P, et al. Hybrid P300-based brain-computer interface to improve usability for people with severe motor disability: electromyographic signals for error correction during a spelling task. *Arch Phys Med Rehabil*. 2015;96(3):S54–61. DOI:10.1016/j.apmr.2014.05.029.
- [137] Riccio A, Leotta F, Bianchi L, et al. Hybrid P300-based brain-computer interface to improve usability for people with severe motor disability: electromyographic signals for error correction during a spelling task. *J Neural Eng*. 2011;8(2):025028. DOI:10.1088/1741-2560/8/2/025028.
- [138] Kübler A, Nijboer F, Kleih S. Hearing the needs of clinical users. *Handb Clin Neurol*. 2020;168:353–368.
- [139] Eidel M, Kübler A. Wheelchair control in a virtual environment by healthy participants using a p300-bci based on tactile stimulation: training effects and usability. *Front Hum Neurosci*. 2020;14:265.
- [140] Ziebell P, Stümpfig J, Eidel M, et al. Stimulus modality influences session-to-session transfer of training effects in auditory and tactile streaming-based P300 brain-computer interfaces. *Sci Rep*. 2020;10(1):11873. DOI:10.1038/s41598-020-67887-6.
- [141] Chavarriaga R, Fried-Oken M, Kleih S, et al. Heading for new shores! Overcoming pitfalls in BCI design. *Brain Comput Interf*(Abingdon, England). 2017;4(1–2):60–73. DOI:10.1080/2326263X.2016.1263916.
- [142] Cunha JD, Perdakis S, Halder S, et al. Post-adaptation effects in a motor imagery brain-computer interface online coadaptive paradigm. *IEEE Access*. 2021;9:41688–41703.
- [143] Roc A, Pillette L, Mladenovic J, et al. A review of user training methods in brain computer interfaces based on mental tasks. *J Neural Eng*. 2020. DOI:10.1088/1741-2552/abca17.
- [144] Halder S, Leinfelder T, Schulz SM, et al. Neural mechanisms of training an auditory event-related potential task in a brain-computer interface context. *Hum Brain Mapp*. 2019;40(8):2399–2412. DOI:10.1002/hbm.24531.
- [145] Blankertz B, Sannelli C, Halder S, et al. Neurophysiological predictor of SMR-based BCI performance. *NeuroImage*. 2010;51(4):1303–1309. DOI:10.1016/j.neuroimage.2010.03.022.
- [146] Hammer A, Vielhaber S, Rodriguez-Fornells A, et al. A neurophysiological analysis of working memory in amyotrophic lateral sclerosis. *Brain Res*. 2011;1421:90–99.
- [147] Formisano R, D'Ippolito M, Riseti M, et al. Vegetative state, minimally conscious state, akinetic mutism and parkinsonism as a continuum of recovery from disorders of consciousness: an exploratory and preliminary study. *Funct Neurol*. 2011;26(1):15–24.
- [148] Giacino JT, Fins JJ, Laureys S, et al. Disorders of consciousness after acquired brain injury: the state of the science. *Nat Rev Neurol*. 2014;10(2):99–114. DOI:10.1038/nrneurol.2013.279.
- [149] Nilsen AS, Juel B, Thüer B, et al. Proposed EEG measures of consciousness: a systematic, comparative review. *PsyArXiv*; 2020.
- [150] Halder S, Juel BE, Nilsen AS, et al. Changes in measures of consciousness during anaesthesia of one hemisphere (wada test). *NeuroImage*. 2021;226:117566.
- [151] Kathner I, Wriessnegger SC, Müller-Putz GR, et al. Effects of mental workload and fatigue on the P300, alpha and theta band power during operation of an ERP (P300) brain-computer interface. *Biol Psychol*. 2014;102:118–129.
- [152] Kleih SC, Nijboer F, Halder S, et al. Motivation modulates the P300 amplitude during brain-computer interface use. *Clin Neurophysiol*. 2010;121(7):1023–1031. DOI:10.1016/j.clinph.2010.01.034.
- [153] Han Y, Valentini E, Halder S. Prediction of tonic pain using support vector machines with phase-based connectivity features. 2021.
- [154] Orlandi S, House SC, Karlsson P, et al. Brain-computer interfaces for children with complex communication needs and limited mobility: a systematic review. *Front Hum Neurosci*. 2021;15:643294.
- [155] Kinney-Lang E, Auyeung B, Escudero J. Expanding the (kaleido)scope: exploring current literature trends for translating electroencephalography (EEG) based brain-computer interfaces for motor rehabilitation in children. *J Neural Eng*. 2016;13(6):061002.
- [156] Fouillen M, Maby E, Le Carrer L, et al. Erp-based BCI training for children with ADHD: motivations and trial design. *GBCIC*; 2017.

- [157] Mikołajewska E, Mikołajewski D. The prospects of brain — computer interface applications in children. *Open Med.* **2014**;9(1):74–79.
- [158] Lim CG, Lee TS, Guan C, et al. A brain-computer interface based attention training program for treating attention deficit hyperactivity disorder. *PloS One.* **2012**;7(10):e46692. DOI:[10.1371/journal.pone.0046692](https://doi.org/10.1371/journal.pone.0046692).
- [159] Norton JJS, Mullins J, Alitz BE, et al. The performance of 9–11-year-old children using an SSVEP-based BCI for target selection. *J Neural Eng.* **2018**;15(5):056012. DOI:[10.1088/1741-2552/aacfd](https://doi.org/10.1088/1741-2552/aacfd).
- [160] House SC. Evaluation of a motor imagery electroencephalography brain-computer interface as an access technology for children (Doctoral dissertation). Toronto Canada: University of Toronto; **2018**.
- [161] Jadavji Z, Zhang J, Paffrath B, et al. Can children with perinatal stroke use a simple brain computer interface?. *Stroke.* **2021**;52(7):2363–2370. DOI:[10.1161/STROKEAHA.120.030596](https://doi.org/10.1161/STROKEAHA.120.030596).
- [162] Kinney-Lang E, Kelly D, Floreani ED, et al. Advancing brain-computer interface applications for severely disabled children through a multidisciplinary national network: summary of the inaugural pediatric bci canada meeting. *Front Hum Neurosci.* **2020**;14(530). DOI:[10.3389/fnhum.2020.593883](https://doi.org/10.3389/fnhum.2020.593883).
- [163] Kelly D, Jadavji Z, Zewdie E, et al. A child's right to play: results from the brain-computer interface game jam 2019 (calgary competition). in 42nd Annual International Conference of the IEEE Engineering in Medicine & Biology Society (EMBC); **2020**.
- [164] Kinney-Lang E, Murji S, Kelly D et al. Designing a flexible tool for rapid implementation of brain-computer interfaces (BCI) in game development. *Annu Int Conf IEEE Eng Med Biol Soc.* **2020**;2020:6078–6081.
- [165] Mullins JL. SSVEP-based BCI performance in children. **2015**.
- [166] Park K, Kihl T, Park S, et al. Fairy tale directed game-based training system for children with ADHD using BCI and motion sensing technologies. *Behaviour Inf Technol.* **2019**;38(6):564–577. DOI:[10.1080/0144929X.2018.1544276](https://doi.org/10.1080/0144929X.2018.1544276).
- [167] Friedrich EV, Suttie N, Sivanathan A, et al. Brain-computer interface game applications for combined neurofeedback and biofeedback treatment for children on the autism spectrum. *Front Neuroeng.* **2014**;7:21.
- [168] Kinney-Lang E, Spyrou L, Ebied A, et al. Tensor-driven extraction of developmental features from varying paediatric EEG datasets. *J Neural Eng.* **2018**;15(4):046024. DOI:[10.1088/1741-2552/aac664](https://doi.org/10.1088/1741-2552/aac664).
- [169] Kinney-Lang E, Ebied A, Escudero J. Building a tensor framework for the analysis and classification of steady-state visual evoked potentials in children. in 2018 26th European Signal Processing Conference (EUSIPCO); **2018**. p. 296–300.
- [170] Forest C, et al. Maturational aspects of visual P300 in children: a research window for pediatric brain computer interface (BCI)*. in 2020 29th IEEE International Conference on Robot and Human Interactive Communication (RO-MAN); **2020**. p. 451–455.
- [171] Kinney-Lang E, Ebied A, Auyeung B, et al. Introducing the joint eeg-development inference (jedi) model: a multi-way, data fusion approach for estimating paediatric developmental scores via EEG. *IEEE Trans Neural Syst Rehabil Eng.* **2019**;27(3):348–357. DOI:[10.1109/TNSRE.2019.2891827](https://doi.org/10.1109/TNSRE.2019.2891827).
- [172] Kinney-lang E, Escudero J. Programming for pediatrics: a literature review of brain-computer interfaces for neurorehabilitation in children. in International Brain-Computer Interface (BCI) Meeting; Asilomar Conference Grounds: Pacific Grove, CA; **2016**.
- [173] Kirton A, Metzler MJ, Craig BT, et al. Perinatal stroke: mapping and modulating developmental plasticity. *Nat Rev Neurol.* **2021**;17(7):415–432. DOI:[10.1038/s41582-021-00503-x](https://doi.org/10.1038/s41582-021-00503-x).
- [174] Letourneau S, Zewdie ET, Jadavji Z, et al. Clinician awareness of brain computer interfaces: a canadian national survey. *J Neuroeng Rehabil.* **2020**;17(1):2. DOI:[10.1186/s12984-019-0624-7](https://doi.org/10.1186/s12984-019-0624-7).
- [175] Europe, S.o.C.P.i., *Surveillance of cerebral palsy in Europe: a collaboration of cerebral palsy surveys and registers*. Surveillance of cerebral palsy in Europe: a collaboration of cerebral palsy surveys and registers. surveillance of cerebral palsy in Europe (SCPE). *Dev Med Child Neurol.* **2000**;42(12):816–824. DOI:[10.1017/S0012162200001511](https://doi.org/10.1017/S0012162200001511).
- [176] Parkes J, Hill N, Platt MJ, et al. Oromotor dysfunction and communication impairments in children with cerebral palsy: a register study. *Dev Med Child Neurol.* **2010**;52(12):1113–1119. DOI:[10.1111/j.1469-8749.2010.03765.x](https://doi.org/10.1111/j.1469-8749.2010.03765.x).
- [177] Cans C, Guillem P, Arnaud C, et al. Prevalence and characteristics of children with cerebral palsy in Europe. *Dev Med Child Neurol.* **2002**;44(9):633–640.
- [178] Kennes J, Rosenbaum P, Hanna SE, et al. Health status of school-aged children with cerebral palsy: information from a population-based sample. *Dev Med Child Neurol.* **2002**;44(4):240–247. DOI:[10.1017/S0012162201002018](https://doi.org/10.1017/S0012162201002018).
- [179] Wood E, Rosenbaum P. The gross motor function classification system for cerebral palsy: a study of reliability and stability over time. *Dev Med Child Neurol.* **2000**;42(5):292–296.
- [180] Alcaide-Aguirre RE, Warschawsky SA, Brown D, et al. Asynchronous brain-computer interface for cognitive assessment in people with cerebral palsy. *J Neural Eng.* **2017**;14(6):066001. DOI:[10.1088/1741-2552/aa7fc4](https://doi.org/10.1088/1741-2552/aa7fc4).
- [181] Daly I, Billinger M, Laparra-Hernández J, et al. On the control of brain-computer interfaces by users with cerebral palsy. *Clin Neurophysiol.* **2013**;124(9):1787–1797. DOI:[10.1016/j.clinph.2013.02.118](https://doi.org/10.1016/j.clinph.2013.02.118).
- [182] Daly I, Scherer R, Billinger M, et al. FORCE: fully online and automated artifact removal for brain-computer interfacing. *IEEE Trans Neural Syst Rehabil Eng.* **2015**;23(5):725–736. DOI:[10.1109/TNSRE.2014.2346621](https://doi.org/10.1109/TNSRE.2014.2346621).
- [183] Scherer R, Billinger M, Wagner J, et al. Thought-based row-column scanning communication board for individuals with cerebral palsy. *Ann Phys Rehabil Med.* **2015**;58(1):14–22. DOI:[10.1016/j.rehab.2014.11.005](https://doi.org/10.1016/j.rehab.2014.11.005).
- [184] Huggins J, Garcia M, Tou S, et al. Comparison of brain-computer interface and eye-gaze interface technology for access to an untimed vocabulary test by people

- with cerebral palsy. in Perth, Australia: Australasian Academy of Cerebral Palsy and Developmental Medicine; 2020.
- [185] Warschausky S, Karlsson P, Huggins J. Brain-computer or eye-gaze interfaces: technical challenges and promise. in AACPDM 73rd Annual and IAACD 2nd Triannual Meeting; Anaheim, California; 2019.
- [186] Kawakami M, Liu M, Otsuka T, et al. Asymmetric skull deformity in children with cerebral palsy: frequency and correlation with postural abnormalities and deformities. *J Rehabil Med.* 2013;45(2):149–153. DOI:10.2340/16501977-1081.
- [187] Minciu I. Clinical correlations in cerebral palsy. *Maedica.* 2012;7(4):319–324.
- [188] Venkateswaran S, Shevell MI. Comorbidities and clinical determinants of outcome in children with spastic quadriplegic cerebral palsy. *Dev Med Child Neurol.* 2008;50(3):216–222.
- [189] Marneweck M, Kuo H-C, Smorenburg ARP, et al. The relationship between hand function and overlapping motor representations of the hands in the contralesional hemisphere in unilateral spastic cerebral palsy. *Neurorehabil Neural Repair.* 2018;32(1):62–72. DOI:10.1177/1545968317745991.
- [190] Tou SLJ, Warschausky SA, Karlsson P, et al. *Individualized electrode subset improves the calibration accuracy of an eeg p300-based brain-computer interface for people with severe cerebral palsy* jne, in revision.
- [191] Himmelmann K, Hagberg G, Wiklund LM, et al. Dyskinetic cerebral palsy: a population-based study of children born between 1991 and 1998. *Dev Med Child Neurol.* 2007;49(4):246–251. DOI:10.1111/j.1469-8749.2007.00246.x.
- [192] Philip SS, Dutton GN. Identifying and characterising cerebral visual impairment in children: a review. *Clin Exp Optom.* 2014;97(3):196–208.
- [193] Scherer R, Schwarz A, Müller-Putz GR, et al. Lets play tic-tac-toe: a brain-computer interface case study in cerebral palsy. in 2016 IEEE International Conference on Systems, Man, and Cybernetics (SMC); 2016.
- [194] Wilson G, Stavisky SD, Willett FR, et al. Decoding spoken English from intracortical electrode arrays in dorsal precentral gyrus. *J Neural Eng.* 2020;17(6):66007. DOI:10.1088/1741-2552/abbfef.
- [195] Stavisky SD, Willett FR, Wilson GH, et al. Neural ensemble dynamics in dorsal motor cortex during speech in people with paralysis. *eLife.* 2019;8:e46015.
- [196] Willett FR, Avansino DT, Hochberg LR, et al. High-performance brain-to-text communication via handwriting. *Nature.* 2021;593(7858):249–254. DOI:10.1038/s41586-021-03506-2.
- [197] Heelan C, Lee J, O'Shea R, et al. Decoding speech from spike-based neural population recordings in secondary auditory cortex of non-human primates. *Commun Biol.* 2019;2(1):466. DOI:10.1038/s42003-019-0707-9.
- [198] Herff C, Krusienski DJ, Kubben P. The potential of stereotactic-eeg for brain-computer interfaces: current progress and future directions. *Front Neurosci.* 2020;14 (February):1–8.
- [199] Gearing M, Kennedy P. Histological confirmation of myelinated neural filaments within the tip of the neurotrophic electrode after a decade of neural recordings. In: *Frontiers in human neuroscience.* 2020; 14.
- [200] Bartels J, Andreassen D, Ehirim P, et al. Neurotrophic electrode: method of assembly and implantation into human motor speech cortex. *J Neurosci Methods.* 2008;174(2):168–176. DOI:10.1016/j.jneumeth.2008.06.030.
- [201] Herff C, Diener L, Angrick M, et al. Generating natural, intelligible speech from brain activity in motor, premotor, and inferior frontal cortices. *Front Neurosci.* 2019;13(November):1–11. DOI:10.3389/fnins.2019.01267.
- [202] Makin JG, Moses DA, Chang EF. Machine translation of cortical activity to text with an encoder–decoder framework. *Nat Neurosci.* 2020;23(4):575–582.
- [203] Herff C, Heger D, de Pestiers A, et al. Brain-to-text: decoding spoken phrases from phone representations in the brain. *Front Neurosci.* 2015;9:217.
- [204] Moses DA, Leonard MK, Makin JG, et al. Real-time decoding of question-and-answer speech dialogue using human cortical activity. *Nat Commun.* 2019;10(1):3096. DOI:10.1038/s41467-019-10994-4.
- [205] Moses DA, Metzger SL, Liu JR, et al. Neuroprosthesis for decoding speech in a paralyzed person with anarthria. *N Engl J Med.* 2021;385(3):217–227. DOI:10.1056/NEJMoa2027540.
- [206] Angrick M, Ottenhoff M, Diener L, et al. Real-time synthesis of imagined speech processes from minimally invasive recordings of neural activity. 2020;(2).
- [207] Brumberg JS, Pitt KM, Burnison JD. A noninvasive brain-computer interface for real-time speech synthesis: the importance of multimodal feedback. *IEEE Trans Neural Syst Rehabil Eng.* 2018;26(4):874–881.
- [208] Berezutskaya J, Vansteensel MJ, Aarnoutse EJ, et al. Open multimodal iEEG-fMRI dataset from naturalistic stimulation with a short audiovisual film. *bioRxiv.* 2021. 2021.06.09.447733.
- [209] Aubinet C, Cassol H, Bodart O, et al. Simplified evaluation of CONsciousness disorders (SECONDS) in individuals with severe brain injury: a validation study. *Ann Phys Rehabil Med.* 2021;64(5):101432. DOI:10.1016/j.rehab.2020.09.001.
- [210] Sanz LRD, Thibaut A, Edlow BL, et al. Update on neuroimaging in disorders of consciousness. *Curr Opin Neurol.* 2021;34(4):488–496. DOI:10.1097/WCO.0000000000000951.
- [211] Thibaut A, Panda R, Annen J et al. Preservation of brain activity in unresponsive patients identifies mcs star. *Ann Neurol.* 2021;90(1):89–100. DOI:10.1002/ana.26095.
- [212] Annen J, Laureys S, Gosseries O. Brain-computer interfaces for consciousness assessment and communication in severely brain-injured patients. *Handb Clin Neurol.* 2020;168:137–152.
- [213] Annen J, Mertel I, Xu R, et al. Auditory and somatosensory p3 are complementary for the assessment of patients with disorders of consciousness. *Brain Sci.* 2020;10(10):748. DOI:10.3390/brainsci10100748.

- [214] Coyle D, Stow J, McCreadie K, et al. Sensorimotor modulation assessment and brain-computer interface training in disorders of consciousness. *Arch Phys Med Rehabil.* 2015;96(3):S62–70. DOI:10.1016/j.apmr.2014.08.024.
- [215] Dayan N, Bigirimana AD, Mccann AE, et al. Towards answering questions in disorders of consciousness and locked-in syndrome with a smr-bci. In 8th Graz Brain-Computer Interface Conference 2019: Bridging Science and Application; Graz: Austria; 2019.
- [216] Guger C, Spataro R, Allison BZ, et al. Complete locked-in and locked-in patients: command following assessment and communication with vibro-tactile p300 and motor imagery brain-computer interface tools. *Front Neurosci.* 2017;11:251.
- [217] Guger C, Spataro R, Pellas F, et al. Assessing command-following and communication with vibro-tactile p300 brain-computer interface tools in patients with unresponsive wakefulness syndrome. *Front Neurosci.* 2018;12:423.
- [218] Murovec N, Heilinger A, Xu R, et al. Effects of a vibro-tactile p300 based brain-computer interface on the coma recovery scale-revised in patients with disorders of consciousness. *Front Neurosci.* 2020;14:294.
- [219] Spataro R, Heilinger A, Allison B, et al. Preserved somatosensory discrimination predicts consciousness recovery in unresponsive wakefulness syndrome. *Clin Neurophysiol.* 2018;129(6):1130–1136. DOI:10.1016/j.clinph.2018.02.131.
- [220] Annen J, Blandiaux S, Lejeune N, et al. BCI performance and brain metabolism profile in severely brain-injured patients without response to command at bedside. *Front Neurosci.* 2018;12:370.
- [221] Thibaut A, Schiff N, Giacino J, et al. Therapeutic interventions in patients with prolonged disorders of consciousness. *Lancet Neurol.* 2019;18(6):600–614. DOI:10.1016/S1474-4422(19)30031-6.
- [222] Bowen A, McKenna K, Tallis RC. Reasons for variability in the reported rate of occurrence of unilateral spatial neglect after stroke. *Stroke.* 1999;30(6):1196–1202.
- [223] Raffin E, Hummel FC. Restoring motor functions after stroke: multiple approaches and opportunities. *Neuroscientist.* 2018;24(4):400–416.
- [224] Remsik AB, Dodd K, Williams L, et al. Behavioral outcomes following brain-computer interface intervention for upper extremity rehabilitation in stroke: a randomized controlled trial. *Front Neurosci.* 2018;12(NOV). 10.3389/fnins.2018.00752.
- [225] Biasiucci A, et al. Brain-actuated functional electrical stimulation elicits lasting arm motor recovery after stroke. *Nature Communications;* 2018.
- [226] Grosse-Wentrup M, Mattia D, Oweiss K. Using brain-computer interfaces to induce neural plasticity and restore function. *J Neural Eng.* 2011;8(2):025004.
- [227] Sebastián-Romagosa M, Cho W, Ortner R, et al. Brain computer interface treatment for motor rehabilitation of upper extremity of stroke patients—a feasibility study. *Front Neurosci.* 2020;14:1056.
- [228] Murovec N, Sebastian-Romagosa M, Dangl S, et al. Preliminary results of a brain-computer interface system based on functional electrical stimulation and avatar feedback for lower extremity rehabilitation of chronic stroke patients. *IEEE Transactions on Systems, Man, and Cybernetics: Systems;* 2020. 2020-October. p. 7–11.
- [229] Rupp R, Rohm M, Schneiders M, et al. Functional rehabilitation of the paralyzed upper extremity after spinal cord injury by noninvasive hybrid neuroprosthesis. *Proc IEEE.* 2015;103(6):954–968. DOI:10.1109/JPROC.2015.2395253.
- [230] Müller-Putz G, Rupp R. Neuroprosthetics and brain-computer interfaces. In: *Spinal cord injury: a guide for clinicians and end users.* Springer Nature; 2021.
- [231] Collinger JL, Wodlinger B, Downey JE, et al. High-performance neuroprosthetic control by an individual with tetraplegia. *Lancet.* 2013;381(9866):557–564. DOI:10.1016/S0140-6736(12)61816-9.
- [232] Hochberg LR, Serruya MD, Friehs GM, et al. Neuronal ensemble control of prosthetic devices by a human with tetraplegia. *see comment. Nature.* 2006;442(7099):164–171. DOI:10.1038/nature04970.
- [233] Pfurtscheller G, Müller GR, Pfurtscheller J, et al. ‘Thought’-control of functional electrical stimulation to restore hand grasp in a patient with tetraplegia. *Neurosci Lett.* 2003;351(1):33–36. DOI:10.1016/S0304-3940(03)00947-9.
- [234] Müller-Putz GR, Scherer R, Pfurtscheller G, et al. EEG-based neuroprosthesis control: a step towards clinical practice. *Neurosci Lett.* 2005;382(1–2):169–174. DOI:10.1016/j.neulet.2005.03.021.
- [235] Rohm M, Schneiders M, Müller C, et al. Hybrid brain-computer interfaces and hybrid neuroprostheses for restoration of upper limb functions in individuals with high-level spinal cord injury. *Artif Intell Med.* 2013;59(2):133–142. DOI:10.1016/j.artmed.2013.07.004.
- [236] Muller-Putz GR, Rupp R, Ofner P, et al. Applying intuitive EEG-controlled grasp neuroprostheses in individuals with spinal cord injury: preliminary results from the MoreGrasp clinical feasibility study. *Annu Int Conf IEEE Eng Med Biol Soc.* 2019;2019:5949–5955.
- [237] Ofner P, Schwarz A, Pereira J, et al. Attempted arm and hand movements can be decoded from low-frequency eeg from persons with spinal cord injury. *Sci Rep.* 2019;9(1):7134. DOI:10.1038/s41598-019-43594-9.
- [238] Ofner P, Schwarz A, Pereira J, et al. Upper limb movements can be decoded from the time-domain of low-frequency EEG. *PLoS One.* 2017;12(8):e0182578. DOI:10.1371/journal.pone.0182578.
- [239] Schwarz A, Ofner P, Pereira J, et al. Decoding natural reach-and-grasp actions from human EEG. *J Neural Eng.* 2018;15(1):016005. DOI:10.1088/1741-2552/aa8911.
- [240] Schwarz A, Pereira J, Kobler R, et al. Unimanual and bimanual reach-and-grasp actions can be decoded from human EEG. *IEEE Trans Biomed Eng.* 2020;67(6):1684–1695. DOI:10.1109/TBME.2019.2942974.
- [241] Sburlea AI, Müller-Putz GR. Exploring representations of human grasping in neural, muscle and kinematic signals. *Sci Rep.* 2018;8(1):16669.

- [242] Sburlea AI, Wilding M, Müller-Putz GR. Disentangling human grasping type from the object's intrinsic properties using low-frequency EEG signals. *Neuroimage: Rep.* **2021**;1(2):100012.
- [243] Martinez-Cagigal V, Kobler RJ, Mondini V, et al. Non-linear online low-frequency EEG decoding of arm movements during a pursuit tracking task. *Annu Int Conf IEEE Eng Med Biol Soc.* **2020**;2020:2981–2985.
- [244] Mondini V, Kobler RJ, Sburlea AI, et al. Continuous low-frequency EEG decoding of arm movement for closed-loop, natural control of a robotic arm. *J Neural Eng.* **2020**;17(4):046031. DOI:[10.1088/1741-2552/aba6f7](https://doi.org/10.1088/1741-2552/aba6f7).
- [245] Müller-Putz GR, Mondini V, Martínez-Cagigal V, et al. Decoding of continuous movement attempt in 2-dimensions from non-invasive low frequency brain signals. In *2021 10th International IEEE/EMBS Conference on Neural Engineering (NER)*; **2021**.
- [246] Kobler RJ, Sburlea AI, Müller-Putz GR. Tuning characteristics of low-frequency EEG to positions and velocities in visuomotor and oculomotor tracking tasks. *Sci Rep.* **2018**;8(1):17713.
- [247] Kobler RJ, Kolesnichenko E, Sburlea AI, et al. Distinct cortical networks for hand movement initiation and directional processing: an EEG study. *Neuroimage.* **2020**;220:117076.
- [248] Hammer J, Fischer J, Ruescher J, et al. The role of ECoG magnitude and phase in decoding position, velocity, and acceleration during continuous motor behavior. *Front Neurosci.* **2013**;7:200.
- [249] Hammer J, Pistohl T, Fischer J, et al. Predominance of movement speed over direction in neuronal population signals of motor cortex: intracranial eeg data and a simple explanatory model. *Cereb Cortex.* **2016**;26(6):2863–2881. DOI:[10.1093/cercor/bhw033](https://doi.org/10.1093/cercor/bhw033).
- [250] Kobler RJ, Sburlea AI, Mondini V, et al. Distance- and speed-informed kinematics decoding improves M/EEG based upper-limb movement decoder accuracy. *J Neural Eng.* **2020**;17(5):056027. DOI:[10.1088/1741-2552/abb3b3](https://doi.org/10.1088/1741-2552/abb3b3).
- [251] Kobler RJ, Sburlea AI, Lopes-Dias C, et al. Corneo-retinal-dipole and eyelid-related eye artifacts can be corrected offline and online in electroencephalographic and magnetoencephalographic signals. *Neuroimage.* **2020**;218:117000.
- [252] Korik A, Sosnik R, Siddique N, et al. *Decoding imagined 3d hand movement trajectories from eeg: evidence to support the use of mu, beta, and low gamma oscillations.* *Front Neurosci.* **2018**;12:130.
- [253] Korik A, Sosnik R, Siddique N, et al. Decoding imagined 3d arm movement trajectories from eeg to control two virtual arms-a pilot study. *Front Neurobot.* **2019**;13:94.
- [254] Lennon O, Tonellato M, Del Felice A, et al. A systematic review establishing the current state-of-the-art, the limitations, and the desired checklist in studies of direct neural interfacing with robotic gait devices in stroke rehabilitation. *Front Neurosci.* **2020**;14:578.
- [255] Mane R, Chouhan T, Guan C. BCI for stroke rehabilitation: motor and beyond. *J Neural Eng.* **2020**;17(4):041001.
- [256] Mane R, Ang KK, Guan C. Brain-Computer interface for stroke rehabilitation. In: N.v. T, editor. *Handbook of Neuroengineering.* Singapore: Springer; **2021**.
- [257] Kwon OY, Lee M-H, Guan C, et al. Subject-independent brain-computer interfaces based on deep convolutional neural networks. *IEEE Trans Neural Netw Learn Syst.* **2020**;31(10):3839–3852. DOI:[10.1109/TNNLS.2019.2946869](https://doi.org/10.1109/TNNLS.2019.2946869).
- [258] Zhang K, Robinson N, Lee S-W, et al. Adaptive transfer learning for EEG motor imagery classification with deep convolutional neural network. *Neural Netw.* **2021**;136:1–10.
- [259] Kuhner D, Fiederer LDJ, Aldinger J, et al. Deep learning based bci control of a robotic service assistant using intelligent goal formulation. **2018**.
- [260] Hartmann KG, Schirrmeister RT, Ball T. Hierarchical internal representation of spectral features in deep convolutional networks trained for eeg decoding. In *International Conference on Brain-Computer Interface (BCI); Asilomar Conference Grounds: Pacific Grove, CA, USA*; **2018**.
- [261] Wodlinger B, Downey JE, Tyler-Kabara EC, et al. Ten-dimensional anthropomorphic arm control in a human brain-machine interface: difficulties, solutions, and limitations. *J Neural Eng.* **2015**;12(1):016011. 2560/12/1/016011 2560/12/1/016011. DOI:[10.1088/1741-2560/12/1/016011](https://doi.org/10.1088/1741-2560/12/1/016011).
- [262] Flesher SN, Collinger JL, Foldes ST, et al. Intracortical microstimulation of human somatosensory cortex. *Sci Transl Med.* **2016**;8(361):361ra141–361ra141. DOI:[10.1126/scitranslmed.aaf8083](https://doi.org/10.1126/scitranslmed.aaf8083).
- [263] Salas MA, Bashford L, Kellis S, et al. Proprioceptive and cutaneous sensations in humans elicited by intracortical microstimulation. *eLife.* **2018**;7:e32904.
- [264] Hiremath SV, Tyler-Kabara EC, Wheeler JJ, et al. Human perception of electrical stimulation on the surface of somatosensory cortex. *PloS One.* **2017**;12(5):e0176020. DOI:[10.1371/journal.pone.0176020](https://doi.org/10.1371/journal.pone.0176020).
- [265] Lee B, Kramer D, Armenta Salas M, et al. Engineering artificial somatosensation through cortical stimulation in humans. *Front Syst Neurosci.* **2018**;12:24.
- [266] Johnson LA, Wander JD, Sarma D, et al. Direct electrical stimulation of the somatosensory cortex in humans using electrocorticography electrodes: a qualitative and quantitative report. *J Neural Eng.* **2013**;10(3):036021. DOI:[10.1088/1741-2560/10/3/036021](https://doi.org/10.1088/1741-2560/10/3/036021).
- [267] Raspopovic S, Capogrosso M, Petrini FM, et al. Restoring natural sensory feedback in real-time bidirectional hand prostheses. *Sci Transl Med.* **2014**;6(222):222ra19. DOI:[10.1126/scitranslmed.3006820](https://doi.org/10.1126/scitranslmed.3006820).
- [268] Davis TS, Wark HAC, Hutchinson DT, et al. Restoring motor control and sensory feedback in people with upper extremity amputations using arrays of 96 micro-electrodes implanted in the median and ulnar nerves. *J Neural Eng.* **2016**;13(3):036001. DOI:[10.1088/1741-2560/13/3/036001](https://doi.org/10.1088/1741-2560/13/3/036001).
- [269] Tan DW, Schiefer MA, Keith MW, et al. A neural interface provides long-term stable natural touch perception. *Sci Transl Med.* **2014**;6(257):257ra138–257ra138. DOI:[10.1126/scitranslmed.3008669](https://doi.org/10.1126/scitranslmed.3008669).
- [270] Chandrasekaran S, Naniavadekar AC, McKernan G, et al. Sensory restoration by epidural stimulation of the lateral spinal cord in upper-limb amputees. *eLife.* **2020**;9. DOI:[10.7554/eLife.54349](https://doi.org/10.7554/eLife.54349).

- [271] Johansson RS, Flanagan JR. Coding and use of tactile signals from the fingertips in object manipulation tasks. *Nat Rev Neurosci.* 2009;10(5):345–359.
- [272] Saal HP, Delhay BP, Rayhaun BC, et al. Simulating tactile signals from the whole hand with millisecond precision. *Proc Natl Acad Sci U S A.* 2017;114(28):E5693–E5702. DOI:10.1073/pnas.1704856114.
- [273] Okorokova EV, He Q, Bensmaia SJ. Biomimetic encoding model for restoring touch in bionic hands through a nerve interface. *J Neural Eng.* 2018;15(6):066033.
- [274] Callier T, Suresh AK, Bensmaia SJ. Neural coding of contact events in somatosensory cortex. *Cereb Cortex.* 2019;29(11):4613–4627. (New York, NY: 1991). DOI:10.1093/cercor/bhy337.
- [275] Kumaravelu K, Tomlinson T, Callier T, et al. A comprehensive model-based framework for optimal design of biomimetic patterns of electrical stimulation for prosthetic sensation. *J Neural Eng.* 2020;17(4):046045. DOI:10.1088/1741-2552/abacd8.
- [276] Oddo CM, Raspopovic S, Artoni F, et al. Intraneural stimulation elicits discrimination of textural features by artificial fingertip in intact and amputee humans. *eLife.* 2016;5:e09148.
- [277] Valle G, Mazzoni A, Iberite F, et al. Biomimetic intra-neural sensory feedback enhances sensation naturalness, tactile sensitivity, and manual dexterity in a bidirectional prosthesis. *Neuron.* 2018;100(1):37–45. e7. DOI:10.1016/j.neuron.2018.08.033.
- [278] Bashford L, Rosenthal I, Kellis S, et al. The neurophysiological representation of imagined somatosensory percepts in human cortex. *The Journal of Neuroscience.* 2021;41(10):JN-RM-2460-20. DOI:10.1523/JNEUROSCI.2460-20.2021.
- [279] Sellers EW, Vaughan TM, Wolpaw JR. A brain-computer interface for long-term independent home use. *Amyotroph Lateral Scler.* 2010;11(5):449–455.
- [280] Leeb R, Perdakis S, Tonin L, et al. Transferring brain-computer interfaces beyond the laboratory: successful application control for motor-disabled users. *Artif Intell Med.* 2013;59(2):121–132. DOI:10.1016/j.artmed.2013.08.004.
- [281] Holz EM, Loic Botrel L, Kübler A. Independent home use of brain painting improves quality of life of two artists in the locked-in state diagnosed with amyotrophic lateral sclerosis. *Brain-Comput Interfaces.* 2015;2(2–3):117–134.
- [282] Wolpaw JR, Bedlack RS, Reda DJ, et al. Independent home use of a brain-computer interface by people with amyotrophic lateral sclerosis. *Neurology.* 2018e258-e267; 913: 10.1212/WNL.0000000000005812
- [283] Pels EGM, Aarnoutse EJ, Leinders S, et al. Stability of a chronic implanted brain-computer interface in late-stage amyotrophic lateral sclerosis. *Clin Neurophysiol.* 2019;130(10):1798–1803. DOI:10.1016/j.clinph.2019.07.020.
- [284] Geronimo A, Simmons Z. TeleBCI: remote user training, monitoring, and communication with an evoked-potential brain-computer interface. *Brain Comput Interfaces.* 2020;7(3–4):57–69.
- [285] Oxley TJ, Yoo PE, Rind GS, et al. Motor neuroprosthesis implanted with neurointerventional surgery improves capacity for activities of daily living tasks in severe paralysis: first in-human experience. *J Neurointerv Surg.* 2021;13(2):102–108. DOI:10.1136/neurintsurg-2020-016862.
- [286] Simeral JD, Hosman T, Saab J, et al. Home use of a percutaneous wireless intracortical brain-computer interface by individuals with tetraplegia. *IEEE Trans Biomed Eng.* 2021;68(7):2313–2325. DOI:10.1109/TBME.2021.3069119.
- [287] Kübler A, Holz EM, Riccio A, et al. The user-centered design as novel perspective for evaluating the usability of BCI-controlled applications. *PloS One.* 2014;9(12):e112392. DOI:10.1371/journal.pone.0112392.
- [288] Pels EGM, Aarnoutse EJ, Ramsey NF, et al. Estimated prevalence of the target population for brain-computer interface neurotechnology in the Netherlands. *Neurorehabil Neural Repair.* 2017;31(7):677–685. DOI:10.1177/1545968317714577.
- [289] Linse K, Aust E, Joos M, et al. Communication matters-pitfalls and promise of hightech communication devices in palliative care of severely physically disabled patients with amyotrophic lateral sclerosis. *Front Neurol.* 2018;9:603.
- [290] Pitt KM, Brumberg JS. Guidelines for feature matching assessment of brain-computer interfaces for augmentative and alternative communication. *Am J Speech Lang Pathol.* 2018;27(3):950–964.
- [291] Pitt KM, Brumberg JS. Evaluating person-centered factors associated with brain-computer interface access to a commercial augmentative and alternative communication paradigm. *Assistive technology: the official journal of RESNA.* 2021;1–10.
- [292] Standards roadmap: neurotechnologies for Brain-Machine Interfacing IEEE Standards Association. 2020
- [293] Freudenburg ZV, Branco MP, Leinders S, et al. Sensorimotor ECoG signal features for bci control: a comparison between people with locked-in syndrome and able-bodied controls. *Front Neurosci.* 2019;13:1058.
- [294] Geronimo A, Stephens HE, Schiff SJ, et al. Acceptance of brain-computer interfaces in amyotrophic lateral sclerosis. *Amyotroph Lateral Scler Frontotemporal Degener.* 2015;16(3–4):258–264. DOI:10.3109/21678421.2014.969275.
- [295] Hill K, Kovacs T, Shin S. Reliability of brain-computer interface language sample transcription procedures. *J Rehabil Res Dev.* 2014;51(4):579–590.
- [296] Hill K, Kovacs T, Shin S. Critical issues using brain-computer interfaces for augmentative and alternative communication. *Arch Phys Med Rehabil.* 2015;96(3):S8–15.
- [297] Hill K, Romich B. A rate index for augmentative and alternative communication. *Int J Speech Technol.* 2002;5:57–64.
- [298] Lenker JA, Koester HH, Smith RO. Toward a national system of assistive technology outcomes measurement. *Assist Technol.* 2021;33(1):1–8.
- [299] McFarland DJ, Sarnacki WA, Wolpaw JR. Brain-computer interface (BCI) operation: optimizing information transfer rates. *Biol Psychol.* 2003;63(3):237–251.

- [300] Pasley BN, David SV, Mesgarani N, et al. Reconstructing speech from human auditory cortex. *PLoS Biol.* **2012**;10(1):e1001251. DOI:[10.1371/journal.pbio.1001251](https://doi.org/10.1371/journal.pbio.1001251).
- [301] Sadeghi S, Maleki A. Accurate estimation of information transfer rate based on symbol occurrence probability in brain-computer interfaces. *Biomed Signal Process Control.* **2019**;54(44):101607.
- [302] Robbins RA, Simmons Z, Bremer BA, et al. Quality of life in ALS is maintained as physical function declines. *Neurology.* **2001**;56(4):442–444. DOI:[10.1212/WNL.56.4.442](https://doi.org/10.1212/WNL.56.4.442).
- [303] Chavarriaga R, Carey C, Luis Contreras-vidal J, et al. Standardization of neurotechnology for brain-machine interfacing: state of the art and recommendations. *IEEE Open J Eng Med Biol.* **2021**;2:71–73.
- [304] Paek AY, Brantley JA, Sujatha Ravindran A, et al. A roadmap towards standards for neurally controlled end effectors. *IEEE Open J Eng Med Biol.* **2021**;2:84–90.
- [305] Initiative IB. Future Neural Therapeutics: technology roadmap white paper version 2. IEEE; **2020**.
- [306] Paek AY, Brantley JA, Evans BJ, et al. Concerns in the blurred divisions between medical and consumer neurotechnology. *IEEE Syst J.* **2021**;15(2):3069–3080. DOI:[10.1109/JSYST.2020.3032609](https://doi.org/10.1109/JSYST.2020.3032609).
- [307] Wexler A, Reiner PB. Oversight of direct-to-consumer neurotechnologies. *Science.* **2019**;363(6424):234–235.
- [308] Wexler A. Separating neuroethics from neurohype. *Nat Biotechnol.* **2019**;37(9):988–990.
- [309] Marchetti M, Priftis K. Brain-computer interfaces in amyotrophic lateral sclerosis: a metanalysis. *Clin Neurophysiol.* **2015**;126(6):1255–1263.
- [310] Wierzgała P, Zapala D, Wojcik GM, et al. Most popular signal processing methods in motor-imagery BCI: a review and meta-analysis. In: *Frontiers in Neuroinformatics.* **2018**. p. 12.
- [311] Eiber C, Delbeke J, Cardoso J, et al. Preliminary minimum reporting requirements for in-vivo neural interface research: i. implantable neural interfaces. *IEEE Open J Eng Med Biol.* **2021**;2:74–83.
- [312] Easttom C, Bianchi L, Valeriani D, et al. A functional model for unifying brain computer interface terminology. *IEEE Open J Eng Med Biol.* **2021**;2:91–96.
- [313] Easttom C. BCI glossary and functional model by the IEEE P2731 working group. *Brain-Comput Interfaces.* **2021**;1–3.
- [314] Zapala D, Hossaini A, Kianpour M, et al. A functional BCI model by the P2731 working group: psychology. *Brain-Comput Interfaces.* **2021**;1–10.
- [315] Bianchi L, Antonietti A, Bajwa G, et al. A functional BCI model by the IEEE P2731 working group: data storage and sharing. *Brain-Comput Interfaces.* **2021**;1–9.
- [316] Eke D, Bernard A, Bjaalie JG, et al. International data governance for neuroscience. *PsyArXiv*; **2021**.
- [317] Ienca M, Haselager P, Emanuel EJ. Brain leaks and consumer neurotechnology. *Nat Biotechnol.* **2018**;36(9):805–810.
- [318] Garden H, Winickoff D, Frahm NM, et al. Responsible innovation in neurotechnology enterprises. *OECD Science, Technology and Industry Working Papers*; **05**. **2019**.
- [319] Hochberg LR, Serruya MD, Friehs GM, et al. Neuronal ensemble control of prosthetic devices by a human with tetraplegia. *Nature.* **2006**;442(7099):164–171.
- [320] Hochberg LR, Bacher D, Jarosiewicz B, et al. Reach and grasp by people with tetraplegia using a neurally controlled robotic arm. *Nature.* **2012**;485(7398):372–375. DOI:[10.1038/nature11076](https://doi.org/10.1038/nature11076).
- [321] Sauter-Starace F, Ratel D, Cretallaz C, et al. Long-term sheep implantation of wimagine(*), a wireless 64-channel electrocorticogram recorder. *Front Neurosci.* **2019**;13:847.
- [322] Larzabal C, Auboiroux V, Karakas S, et al. The riemannian spatial pattern method: mapping and clustering movement imagery using riemannian geometry. *J Neural Eng.* **2021**;18(5):056014. DOI:[10.1088/1741-2552/abf291](https://doi.org/10.1088/1741-2552/abf291).
- [323] Brandman DM, Hosman T, Saab J, et al. Rapid calibration of an intracortical brain-computer interface for people with tetraplegia. *J Neural Eng.* **2018**;15(2):026007. DOI:[10.1088/1741-2552/aa9ee7](https://doi.org/10.1088/1741-2552/aa9ee7).
- [324] Gunduz A, Opri E, Gilron R, et al. Adding wisdom to ‘smart’ bioelectronic systems: a design framework for physiologic control including practical examples. *Bioelectron Med (Lond).* **2019**;2(1):29–41. DOI:[10.2217/bem-2019-0008](https://doi.org/10.2217/bem-2019-0008).
- [325] Allison BZ. Toward ubiquitous BCIs. In: P. G. GB, B A, editors. *Brain-computer interfaces.* Berlin Heidelberg: Springer; **2009**. p. 357–387.
- [326] Burwell S, Sample M, Racine E. Ethical aspects of brain computer interfaces: a scoping review. *BMC Med Ethics.* **2017**;18(1):60.
- [327] Haselager P, Mecacci G, Wolkenstein A, et al. Can BCIs enlighten the concept of agency? A plea for an experimental philosophy of Neurotechnology. In: Friedrich O, editors. *Clinical Neurotechnology meets Artificial Intelligence: philosophical, ethical, legal and social implications.* Cham, Switzerland: Springer Nature; **2021**. p. 55.
- [328] Schneider MJ, Fins JJ, Wolpaw JR. Ethical issues in BCI research. In: Wolpaw JR, and Wolpaw EW, editors. *Brain-Computer Interfaces. Principles and practice.* New York, USA: Oxford University Press; **2012**. p. 373–383.
- [329] Yuste R, Goering S, Arcas BAY, et al. Four ethical priorities for neurotechnologies and AI. *Nature.* **2017**;551(7679):159–163. DOI:[10.1038/551159a](https://doi.org/10.1038/551159a).
- [330] Nijboer F, Clausen J, Allison BZ, et al. The asilomar survey: stakeholders’ opinions on ethical issues related to brain-computer interfacing. *Neuroethics.* **2013**;6(3):541–578. DOI:[10.1007/s12152-011-9132-6](https://doi.org/10.1007/s12152-011-9132-6).
- [331] Pham M, Goering S, Sample M, et al. Asilomar survey: researcher perspectives on ethical principles and guidelines for BCI research. *Brain-Comput Interfaces.* **2018**;5(4):97–111. DOI:[10.1080/2326263X.2018.1530010](https://doi.org/10.1080/2326263X.2018.1530010).
- [332] Sample M, Sattler S, Blain-Moraes S, et al. Do publics share experts’ concerns about brain-computer interfaces? A trinational survey on the ethics of neural technology. *Sci Technol Hum Values.* **2020**;45(6):1242–1270. DOI:[10.1177/0162243919879220](https://doi.org/10.1177/0162243919879220).
- [333] Cinel C, Valeriani D, Poli R. Neurotechnologies for human cognitive augmentation: current state of the art and future prospects. *Front Hum Neurosci.* **2019**;13:13.

- [334] Matran-Fernandez A, Poli R. Brain-computer interfaces for detection and localization of targets in aerial images. *IEEE Trans Biomed Eng.* **2017**;64(4):959–969.
- [335] Matran-Fernandez A, Poli R, Hu D. Towards the automated localisation of targets in rapid image-sifting by collaborative brain-computer interfaces. *PLoS One.* **2017**;12(5):e0178498.
- [336] Poli R, Valeriani D, Cinel C. Collaborative brain-computer interface for aiding decision-making. *PLoS One.* **2014**;9(7):e102693.
- [337] Valeriani D, Cinel C, Poli R. Group augmentation in realistic visual-search decisions via a hybrid brain-computer interface. *Sci Rep.* **2017**;7(1):7772.
- [338] Valeriani D, Poli R, Cinel C. Enhancement of group perception via a collaborative brain-computer interface. *IEEE Trans Biomed Eng.* **2017**;64(6):1238–1248.
- [339] Valeriani D, Poli R, Schwenker F. Cyborg groups enhance face recognition in crowded environments. *PLoS One.* **2019**;14(3):e0212935.
- [340] Curtin A, Ayaz H, Tang Y, et al. Enhancing neural efficiency of cognitive processing speed via training and neurostimulation: an fNIRS and TMS study. *Neuroimage.* **2019**;198:73–82.
- [341] McKendrick R, Falcone B, Scheldrup M, et al. Effects of transcranial direct current stimulation on baseline and slope of prefrontal cortex hemodynamics during a spatial working memory task. *Front Hum Neurosci.* **2020**;14:64.
- [342] Rao RP, Stocco A, Bryan M, et al. A direct brain-to-brain interface in humans. *PLoS One.* **2014**;9(11):e111332. DOI:10.1371/journal.pone.0111332.
- [343] Yang Y, Qiao S, Sani OG, et al. Modelling and prediction of the dynamic responses of large-scale brain networks during direct electrical stimulation. *Nat Biomed Eng.* **2021**;5(4):324–345. DOI:10.1038/s41551-020-00666-w.
- [344] Shanechi MM. Brain-machine interfaces from motor to mood. *Nat Neurosci.* **2019**;22(10):1554–1564.
- [345] Dehais F, Karwowski W, Ayaz H, et al. Brain at work and in everyday life as the next frontier: grand field challenges for neuroergonomics. *Front Neuroergonom.* **2020**;1. DOI:10.3389/fnrgo.2020.583733.
- [346] Neuroergonomics: the brain at work and in everyday life. London: Academic Press; **2018**.
- [347] Kosmyna N, Maes P. AttentivU: an EEG-based closed-loop biofeedback system for real-time monitoring and improvement of engagement for personalized learning. *Sensors (Basel).* **2019**;19(23):5200.
- [348] Van Dijk J, Hummels C. Beyond distributed representation: embodied cognition design supporting socio-sensorimotor couplings. In 8th International Conference on Tangible, Embedded and Embodied Interaction; Munich, Germany; **2014**.
- [349] Frey J, Hachet M, Lotte F. EEG-based neuroergonomics for 3d user interfaces: opportunities and challenges. *Le travail humain.* **2017**;80(1):73–92.
- [350] Jeunet C, Hauw D, Millán JDR. Sport psychology: technologies ahead. *Front Sport Active Liv.* **2020**;2(10). DOI:10.3389/fspor.2020.00010
- [351] Di Fronso S, Fiedler P, Tamburro G, et al. Dry EEG in sports sciences: a fast and reliable tool to assess individual alpha peak frequency changes induced by physical effort. *Front Neurosci.* **2019**;13:982.
- [352] Callan DE, Dehais F. Chapter 9 - neuroergonomics for aviation. In: Ayaz H, Dehais F, editors. *Neuroergonomics*. Academic Press; **2019**. p. 55–58.
- [353] Fairclough SH, Lotte F. Grand challenges in neurotechnology and system neuroergonomics. *Front Neuroergonom.* **2020**;1(2). DOI:10.3389/fnrgo.2020.602504
- [354] Stone DB, Tamburro G, Fiedler P, et al. Automatic removal of physiological artifacts in eeg: the optimized fingerprint method for sports science applications. *Front Hum Neurosci.* **2018**;12(96). DOI:10.3389/fnhum.2018.00096.
- [355] Casso MI, Jeunet C, Roy RN. Heading for motor imagery brain-computer interfaces (MI-BCIs) usable out-of-the-lab: impact of dry electrode setup on classification accuracy. In 2021 10th International IEEE/EMBS Conference on Neural Engineering (NER); **2021**. p. 690–693.
- [356] Dehais F, Duprès A, Blum S, et al. Monitoring pilot's mental workload using erps and spectral power with a six-dry-electrode eeg system in real flight conditions. *Sensors (Basel).* **2019**;19(6):1324. DOI:10.3390/s19061324.
- [357] Mühl C, Heylen D, Nijholt A, et al. *Affective brain-computer interfaces: neuroscientific approaches to affect detection* in *Oxford Handbook of affective computing*. Calvo R, Editors. New York: Oxford University Press; **2015**. p. 217–237.
- [358] Nijholt A. *Brain art. Brain-Computer Interfaces and artistic expression*. Human-computer interaction series. London UK: Springer; **2019**.
- [359] Wadeson A, Nijholt A, Nam CS. Artistic brain-computer interfaces: current state-of-art of control mechanisms. *Brain-Comput Interfaces.* **2015**;2(2–3):70–75.
- [360] Prpa M, Pasquier P. *Brain-Computer Interfaces in contemporary art: a state of the art and taxonomy*. In: Nijholt A, editor. *Brain art. Brain-Computer Interfaces and artistic expression*. London (UK): Springer; **2019**. p. 65–115.
- [361] King JL. Art therapy, trauma, and neuroscience: theoretical and practical perspectives. Routledge; **2016**. p. 234. pages.
- [362] Scott SM, Gehrke L. Neurofeedback during creative expression as a therapeutic tool. In: Contreras-Vidal J, editors. *Mobile brain-body imaging and the neuroscience of art, innovation and creativity*. Cham Switzerland: Springer Nature; **2019**. p. 161–166.
- [363] Scott SM, Raftery C, Anderson C. Advancing the rehabilitative and therapeutic potentials of BCI and noninvasive systems. In: Nijholt A, editor. *Brain art: brain-computer interfaces for artistic expression*. Springer International Publishing: Switzerland; **2019**. p. 327–354.
- [364] Valjamae A, et al. The brainhack project. In *Proceedings of the 2017 ACM Workshop on An Application-oriented Approach to BCI out of the laboratory - BCIforReal '17*; ACM Press; **2017**.
- [365] Guger C, et al. The BR4IN.IO hackathons. In: A Neditors. *Brain art. Brain-Computer Interfaces and artistic expression*. London (UK): Springer; **2019**. p. 447–473.

- [366] Väljamäe A. TLÜ 13: Neurotheatre as a research tool by Aleksander Väljamäe; 2021. <https://www.youtube.com/watch?v=L4xA0vU5XtY>.
- [367] Ramchurn R, et al. Brain-controlled cinema. In: Nijholt A editors. Brain art. Brain-Computer Interfaces and artistic expression. Cham Switzerland: Springer; 2019. p. 377–408.
- [368] Rosenboom D, Mullen T. More than one—artistic explorations with multi-agent BCIs. In: Nijholt A, editor. Brain art. Brain-Computer Interfaces and artistic expression. Cham Switzerland: Springer; 2019. p. 117–143.
- [369] Williams D. *Evaluating BCI for musical expression: historical approaches, challenges and benefits*, in brain art. In: Nijholt A, editor. Brain-computer interfaces and artistic expression. Cham Switzerland: Springer; 2019. p. 145–158.
- [370] Hildt E. Affective brain-computer music interfaces-drivers and implications. Front Hum Neurosci. 2021;15: 711407.
- [371] Pearlman E. Brain opera: exploring surveillance in 360-degree immersive theatre. PAJ: J Perform Art. 2017;39(2):79–85.
- [372] Pearlman E. AI comes of age. PAJ: J Perform Art. 2020;42(3):55–62.
- [373] Pearlman E, Ellen Pearlman RU, Durbas Iela, et al. Is there a place in human consciousness where surveillance cannot go? Noor: a brain opera. Leonardo. 2021;54(5):542–546. DOI:10.1162/leon_a_02008.